



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB106907

Reagent/Standard	Lot/Log #
Residual chlorine std, and CCV , 0.4PPM	WP80771
Residual chlorine ICV-LCS, 0.4PPM	WP80770
Chlorine Calibration std, 0.1ppm	WP80765
Chlorine Calibration std, 0.2ppm	WP80766
Chlorine Calibration std, 0.8ppm	WP80768
Chlorine Calibration std, 0.0ppm	WP80764
Chlorine Calibration std, 1.6ppm	WP80769
Chlorine Calibration std & CCV, 0.4p	
Total Chlorine Powder Pillows	W2249

Intercept: -0.0014

Slope: 1.0060

Regression: 0.999995

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	12/23/2019	14:02
2	CAL2	0.1	1	0.000	0.100	0.100	12/23/2019	14:03
3	CAL3	0.2	1	0.000	0.200	0.200	12/23/2019	14:03
4	CAL4	0.4	1	0.000	0.400	0.400	12/23/2019	14:04
5	CAL5	0.8	1	0.000	0.800	0.800	12/23/2019	14:04
6	CAL6	1.6	1	0.000	1.610	1.610	12/23/2019	14:05



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB106907

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3890	12/23/2019	14:05
2	ICB				1	0.0000	0.0000	0.0000	0.0010	12/23/2019	14:06
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4090	12/23/2019	14:06
4	CCB1				1	0.0000	0.0000	0.0000	0.0010	12/23/2019	14:07
5	LB106907BL	50	50		1	0.0000	0.0000	0.0000	0.0010	12/23/2019	14:08
6	LB106907BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3990	12/23/2019	14:08
7	K6416-01	50	50		1	0.0100	0.0900	0.0800	0.0810	12/23/2019	14:09
8	K6416-01DUP	50	50		1	0.0100	0.0900	0.0800	0.0810	12/23/2019	14:10
9	CCV2			0.4	1	0.0000	0.4000	0.4000	0.3990	12/23/2019	14:11
10	CCB2				1	0.0000	0.0000	0.0000	0.0010	12/23/2019	14:11

WORKLIST(Hardcopy Internal Chain)

UB 106907

WorkList Name : RCL122319 WorkList ID : 134502 Department : Wet-Chemistry Date : 12-23-2019 13:28:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/16/2019	Water	K6416-01	Residual Chlorine	Cool 4 deg C	STER17	JS9600		12/16/2019	SM4500 Cl G

Date/Time 12-23-19 13:45
 Received by: AE
 Relinquished by: CS

Date/Time 12-23-19 14:30
 Received by: CS
 Relinquished by: AE