



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106921

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/24/2019	17:05
2	CAL2	1	Water	NA	NA	20.1	7.00	12/24/2019	17:06
3	CAL3	1	Water	NA	NA	20.1	10.03	12/24/2019	17:07
4	ICV	1	Water	NA	NA	20.2	7.01	12/24/2019	17:09
5	CCV1	1	Water	NA	NA	20.1	2.01	12/24/2019	17:10
6	K6435-01	1	Solid	20.02	20	20.2	9.98	12/24/2019	17:11
7	K6435-01DUP	1	Solid	20.03	20	20.1	9.99	12/24/2019	17:12
8	K6435-02	1	Solid	20.05	20	20.2	10.15	12/24/2019	17:15
9	K6435-04	1	Solid	20.03	20	20.7	10.18	12/24/2019	17:16
10	CCV2	1	Water	NA	NA	20.1	12.02	12/24/2019	17:19

106921

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6437 WorkList ID : 134542 Department : Wet-Chemistry Date : 12-24-2019 16:00:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/23/2019	Solid	K6435-01	pH	Cool 4 deg C	DVIR01	E11	SB-4D-(21.5-23.5)	12/23/2019	9045D
12/23/2019	Solid	K6435-02	pH	Cool 4 deg C	DVIR01	E11	SB-4E-(21.5-23.5)	12/23/2019	9045D
12/23/2019	Solid	K6435-04	pH	Cool 4 deg C	DVIR01	E11	SB-3E-(09.5-11.5)	12/23/2019	9045D

Date/Time 12.24.19 16.00
 Received by: Jg
 Relinquished by: Jg

Date/Time 12.26.19 18.00
 Received by: Jg
 Relinquished by: Jg