

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB106935

ANALYST: Daniel
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP80247
IODINE SOLUTION .025N 1L	W2508
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB125735BL		1	5.04	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	12/26/2019	11:50
2	K6436-08		1	5.02	50	2.00	1.92	3.82	1.90	0.10	0.02	1.59	12/26/2019	11:52
3	K6436-08DUP		1	5.04	50	2.00	3.82	5.70	1.88	0.12	0.04	3.17	12/26/2019	11:54
4	K6436-09		1	5.01	50	2.00	5.70	7.56	1.86	0.14	0.06	4.79	12/26/2019	11:56
5	K6437-12		1	5.00	50	2.00	0.00	1.90	1.90	0.10	0.02	1.60	12/26/2019	12:00
6	K6437-13		1	5.03	50	2.00	1.90	3.78	1.88	0.12	0.04	3.18	12/26/2019	12:03
7	K6437-15		1	5.02	50	2.00	3.78	5.64	1.86	0.14	0.06	4.78	12/26/2019	12:05

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume