


LB106936
INSTRUMENT ID C01

7470A

μ Abs.:

47822



Linear

A= 0.0000e+000

B= 2.1073e-004

C= 8.3647e-003

Rho= 0.9998706

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	24	0.000	24				
0.2	0.200	0.216	0.016	987	0.0 %	987				
2.5	2.500	2.507	0.007	11859	0.0 %	11859				
5.0	5.000	4.992	-0.008	23647	0.0 %	23647				
7.5	7.500	7.385	-0.115	35006	0.0 %	35006				
10.0	10.000	10.086	0.086	47822	0.0 %	47822				

LB106936 INSTRUMENT ID CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Dec 2019 13:43:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	24	-	0.0000	7470A	PPB	26 Dec 2019 13:50:26	S	Std
0.2 - 1	50.2	987	-	0.2000	7470A	PPB	26 Dec 2019 14:12:14	S	Std
2.5 - 1	52.5	11859	-	2.5000	7470A	PPB	26 Dec 2019 14:14:22	S	Std
5.0 - 1	55.0	23647	-	5.0000	7470A	PPB	26 Dec 2019 14:16:30	S	Std
7.5 - 1	57.5	35006	-	7.5000	7470A	PPB	26 Dec 2019 14:25:28	S	Std
10.0 - 1	510.0	47822	-	10.0000	7470A	PPB	26 Dec 2019 14:27:35	S	Std
ICV41 - 1	ICV41	17926	3.7859 ✓	-	7470A	PPB	26 Dec 2019 14:30:06	U	SMPL
ICB41 - 1	ICB41	16	0.0117 ✓	-	7470A	PPB	26 Dec 2019 14:32:12	U	SMPL
CCV75 - 1	CCV75	24199	5.1079 ✓	-	7470A	PPB	26 Dec 2019 14:34:18	U	SMPL
CCB75 - 1	CCB75	12	0.0109 ✓	-	7470A	PPB	26 Dec 2019 14:36:25	U	SMPL
CRA - 1	CRA	934	0.2052 ✓	-	7470A	PPB	26 Dec 2019 14:38:31	U	SMPL
HighStd - 1	HighStd	48016	10.1269 ✓	-	7470A	PPB	26 Dec 2019 14:40:37	U	SMPL
ChkStd - 1	ChkStd	31995	6.7507 ✓	-	7470A	PPB	26 Dec 2019 14:42:44	U	SMPL
PB125741BL - 1	PBW	-17	0.0048 ✓	-	7470A	PPB	26 Dec 2019 14:44:51	U	SMPL
PB125741BS - 1	LCSW	18002	3.8020 ✓	-	7470A	PPB	26 Dec 2019 14:46:58	U	SMPL
K6395-04 - 1	WC	27	0.0141 ✓	-	7470A	PPB	26 Dec 2019 14:49:05	U	SMPL
K6395-04DUP - 1	WCDUP	56	0.0202 ✓	-	7470A	PPB	26 Dec 2019 14:51:12	U	SMPL
K6395-04MS - 1	WCMS	18594	3.9267 ✓	-	7470A	PPB	26 Dec 2019 14:57:32	U	SMPL
K6395-04MSD - 1	WCMSD	18818	3.9739 ✓	-	7470A	PPB	26 Dec 2019 15:01:25	U	SMPL
PB125742BL - 1	PBW	1	0.0086 ✓	-	7470A	PPB	26 Dec 2019 15:03:31	U	SMPL
CCV76 - 1	CCV76	24032	5.0727 ✓	-	7470A	PPB	26 Dec 2019 15:08:02	U	SMPL
CCB76 - 1	CCB76	-7	0.0069 ✓	-	7470A	PPB	26 Dec 2019 15:10:07	U	SMPL
PB125742BS - 1	LCSW	18307	3.8662 ✓	-	7470A	PPB	26 Dec 2019 15:12:14	U	SMPL
K6118-13 - 1	SU-01-122319	54	0.0197 ✓	-	7470A	PPB	26 Dec 2019 15:14:20	U	SMPL
K6118-13DUP - 1	SU-01-122319DUP	77	0.0246 ✓	-	7470A	PPB	26 Dec 2019 15:16:26	U	SMPL
K6118-13MS - 1	SU-01-122319MS	19445	4.1060 ✓	-	7470A	PPB	26 Dec 2019 15:18:33	U	SMPL
K6118-13MSD - 1	SU-01-122319MSD	19452	4.1075 ✓	-	7470A	PPB	26 Dec 2019 15:20:40	U	SMPL
K6413-14 - 1	K005-AA-WC-1	7	0.0098 ✓	-	7470A	PPB	26 Dec 2019 15:22:47	U	SMPL
K6413-15 - 1	K005-AA-WC-2	30	0.0147 ✓	-	7470A	PPB	26 Dec 2019 15:24:54	U	SMPL
K6413-16 - 1	K005-AB-WC-1	53	0.0195 ✓	-	7470A	PPB	26 Dec 2019 15:27:02	U	SMPL
K6414-09 - 1	K183-WC-01	41	0.0170 ✓	-	7470A	PPB	26 Dec 2019 15:29:10	U	SMPL
K6414-10 - 1	K183-WC-02	31	0.0149 ✓	-	7470A	PPB	26 Dec 2019 15:31:16	U	SMPL
CCV77 - 1	CCV77	24180	5.1039 ✓	-	7470A	PPB	26 Dec 2019 15:33:23	U	SMPL
CCB77 - 1	CCB77	3	0.0090 ✓	-	7470A	PPB	26 Dec 2019 15:35:29	U	SMPL
K6436-08 - 1	K009-WC-1	41	0.0170 ✓	-	7470A	PPB	26 Dec 2019 15:37:36	U	SMPL
K6436-09 - 1	K009-WC-2	42	0.0172 ✓	-	7470A	PPB	26 Dec 2019 15:39:42	U	SMPL
PB125743BL - 1	PBW	7	0.0098 ✓	-	7470A	PPB	26 Dec 2019 15:41:48	U	SMPL
PB125743BS - 1	LCSW	19751	4.1705 ✓	-	7470A	PPB	26 Dec 2019 15:43:54	U	SMPL
K6434-07 - 1	GW-2	33	0.0153 ✓	-	7470A	PPB	26 Dec 2019 15:46:01	U	SMPL
K6434-07DUP - 1	GW-2DUP	32	0.0151 ✓	-	7470A	PPB	26 Dec 2019 15:48:09	U	SMPL
K6434-07MS - 1	GW-2MS	20518	4.3322 ✓	-	7470A	PPB	26 Dec 2019 15:50:17	U	SMPL
K6434-07MSD - 1	GW-2MSD	20518	4.3322 ✓	-	7470A	PPB	26 Dec 2019 15:55:41	U	SMPL
K6434-08 - 1	GW-1	32	0.0151 ✓	-	7470A	PPB	26 Dec 2019 15:57:48	U	SMPL
K6434-09 - 1	GW-3	54	0.0197 ✓	-	7470A	PPB	26 Dec 2019 15:59:55	U	SMPL
CCV78 - 1	CCV78	23613	4.9844 ✓	-	7470A	PPB	26 Dec 2019 16:02:02	U	SMPL
CCB78 - 1	CCB78	19	0.0124 ✓	-	7470A	PPB	26 Dec 2019 16:04:08	U	SMPL
K6434-10 - 1	GW-5	45	0.0178 ✓	-	7470A	PPB	26 Dec 2019 16:06:14	U	SMPL
PB125744BL - 1	PBW	36	0.0160 ✓	-	7470A	PPB	26 Dec 2019 16:08:21	U	SMPL
PB125744BS - 1	LCSW	19992	4.2213 ✓	-	7470A	PPB	26 Dec 2019 16:10:27	U	SMPL
K6406-02 - 1	C-WATER_12.20.19	37	0.0162 ✓	-	7470A	PPB	26 Dec 2019 16:12:33	U	SMPL
K6406-02DUP - 1	C-WATER_12.20.19DUP	55	0.0200 ✓	-	7470A	PPB	26 Dec 2019 16:14:40	U	SMPL
K6406-02MS - 1	C-WATER_12.20.19MS	19876	4.1969 ✓	-	7470A	PPB	26 Dec 2019 16:16:48	U	SMPL
K6406-02MSD - 1	C-WATER_12.20.19MSD	20045	4.2325 ✓	-	7470A	PPB	26 Dec 2019 16:18:55	U	SMPL
K6434-01 - 1	GW-2	28	0.0143 ✓	-	7470A	PPB	26 Dec 2019 16:21:03	U	SMPL
K6434-02 - 1	GW-1	39	0.0166 ✓	-	7470A	PPB	26 Dec 2019 16:23:11	U	SMPL
K6434-05 - 1	GW-3	55	0.0200 ✓	-	7470A	PPB	26 Dec 2019 16:25:18	U	SMPL
CCV79 - 1	CCV79	23773	5.0181 ✓	-	7470A	PPB	26 Dec 2019 16:27:25	U	SMPL
CCB79 - 1	CCB79	18	0.0122 ✓	-	7470A	PPB	26 Dec 2019 16:29:32	U	SMPL
K6434-06 - 1	GW-5	50	0.0189 ✓	-	7470A	PPB	26 Dec 2019 16:31:39	U	SMPL
K6395-04LX5 - 1		-4	0.0075 ✓	-	7470A	PPB	26 Dec 2019 16:33:46	U	SMPL
K6118-13LX5 - 1		13	0.0111 ✓	-	7470A	PPB	26 Dec 2019 16:35:53	U	SMPL
K6434-07LX5 - 1		-3	0.0077 ✓	-	7470A	PPB	26 Dec 2019 16:38:00	U	SMPL
K6406-02LX5 - 1		0	0.0084 ✓	-	7470A	PPB	26 Dec 2019 16:40:05	U	SMPL
PB125741TB - 1	PB125651TB	14	0.0113 ✓	-	7470A	PPB	26 Dec 2019 16:42:12	U	SMPL
PB125742TB - 1	PB125688TB	-1	0.0082 ✓	-	7470A	PPB	26 Dec 2019 16:44:19	U	SMPL
CCV80 - 1	CCV80	23673	4.9970 ✓	-	7470A	PPB	26 Dec 2019 16:46:27	U	SMPL
CCB80 - 1	CCB80	-26	0.0029 ✓	-	7470A	PPB	26 Dec 2019 16:48:35	U	SMPL