



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 12/27/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 16:50
In Date: 12/26/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 12/27/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB106942

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
K6113-11	OR-03-122619	1	1.14	9.67	8.85	90.4	
K6113-12	OR-03-122619	2	1.16	9.97	9.46	94.2	
K6113-13	OR-03-122619-1	3	1.18	9.56	8.85	91.5	
K6113-14	OR-03-122619-2	4	1.14	9.77	8.82	89.0	
K6117-10	TR-05-122619	5	1.19	9.87	8.97	89.6	
K6117-11	TR-05-122619	6	1.12	9.66	8.8	89.9	
K6441-01	K321-SB-01	7	1.13	6.92	6.72	96.5	
K6441-02	K321-SB-02	8	1.14	6.78	6.29	91.3	
K6441-03	K321-SB-03	9	1.13	5.54	4.98	87.3	
K6441-04	K321-SB-04	10	1.14	7.01	6.36	88.9	
K6441-05	K321-SB-05	11	1.15	5.9	5.31	87.6	
K6441-06	K321-WC-01	12	1.17	9.68	9.04	92.5	
K6441-07	K321-WC-02	13	1.12	9.82	8.74	87.6	
K6442-01	K615-SB1	14	1.13	6.21	5.73	90.6	
K6442-02	K615-SB2	15	1.12	6.04	5.72	93.5	
K6442-03	K615-SB3	16	1.18	6.01	5.76	94.8	
K6442-04	K615-SB4	17	1.12	6.35	5.93	92.0	
K6442-05	K615-SB5	18	1.14	5.7	5.5	95.6	
K6442-06	K615-WC-1	19	1.14	9.86	9.28	93.3	
K6443-01	PH-1	20	1.14	9.98	9.32	92.5	
K6443-02	PH-1	21	1.17	9.52	8.92	92.8	
K6443-03	PH-3	22	1.19	9.86	9.12	91.5	
K6443-04	PH-3	23	1.15	9.69	9.02	92.2	
K6443-05	PH-2	24	1.19	9.72	9.00	91.6	
K6443-06	PH-2	25	1.14	9.58	9.06	93.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$