



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106957

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/27/2019	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	12/27/2019	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	12/27/2019	17:02
4	ICV	1	Water	NA	NA	20.1	6.99	12/27/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	12/27/2019	17:06
6	K6114-02	1	Solid	20.02	20	21.6	8.08	12/27/2019	17:08
7	K6455-07	1	Solid	20.03	20	21.1	10.76	12/27/2019	17:10
8	K6456-02	1	Solid	20.04	20	21.4	7.84	12/27/2019	17:11
9	K6456-04	1	Solid	20.02	20	21.7	7.87	12/27/2019	17:12
10	K6456-04DUP	1	Solid	20.03	20	21.6	7.88	12/27/2019	17:13
11	CCV2	1	Water	NA	NA	20.1	12.02	12/27/2019	17:15

106957

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K6455 WorkList ID : 134608 Department : Wet-Chemistry Date : 12-27-2019 16:10:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/27/2019	Solid	K6114-02	Corrosivity	Cool 4 deg C	PSEG05	E31	EO-O1-122719	12/27/2019	9045D
12/27/2019	Solid	K6455-07	Corrosivity	Cool 4 deg C	ATCE02	E61	K435-WC-1	12/27/2019	9045D
12/27/2019	Solid	K6456-02	Corrosivity	Cool 4 deg C	PSEG01	E22	SB-03	12/27/2019	9045D
12/27/2019	Solid	K6456-04	Corrosivity	Cool 4 deg C	PSEG01	E22	SB-04	12/27/2019	9045D

Date/Time 12-27-19 16:11
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 12-27-19 17:40
 Received by: [Signature]
 Relinquished by: [Signature]