


LB106964
INSTRUMENT ID CV1

7470A

Linear



A= 0.0000e+000
B= 2.4883e-004
C= -2.2556e-001
Rho= 0.9993899
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.184	-0.184	169	0.000	169				
0.2	0.200	0.190	-0.010	1672	0.0 %	1672				
2.5	2.500	2.704	0.204	11772	0.0 %	11772				
5.0	5.000	5.083	0.083	21333	0.0 %	21333				
7.5	7.500	7.534	0.034	31185	0.0 %	31185				
10.0	10.000	9.872	-0.128	40580	0.0 %	40580				

LB106964 INSTRUMENT ID CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Dec 2019 13:17:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC-0	169	-	0.0000	7470A	PPB	30 Dec 2019 13:30:35	S	Std
0.2 - 1	SC-2	1672	-	0.2000	7470A	PPB	30 Dec 2019 13:38:29	S	Std
2.5 - 1	SR5	11772	-	2.5000	7470A	PPB	30 Dec 2019 13:40:37	S	Std
5.0 - 1	SG-0	21333	-	5.0000	7470A	PPB	30 Dec 2019 13:42:45	S	Std
7.5 - 1	ST-5	31185	-	7.5000	7470A	PPB	30 Dec 2019 13:48:56	S	Std
10.0 - 1	SI-0-0	40580	-	10.0000	7470A	PPB	30 Dec 2019 13:53:38	S	Std
ICV43 - 1	ICV43	16040	3.7657 ✓	-	7470A	PPB	30 Dec 2019 13:56:01	U	SMPL
ICB43 - 1	ICB43	151	-0.1880 ✓	-	7470A	PPB	30 Dec 2019 13:58:07	U	SMPL
CCV83 - 1	CCV83	21365	5.0908 ✓	-	7470A	PPB	30 Dec 2019 14:01:38	U	SMPL
CCB83 - 1	CCB83	234	-0.1673 ✓	-	7470A	PPB	30 Dec 2019 14:03:44	U	SMPL
CRA - 1	CRA	1683	0.1932 ✓	-	7470A	PPB	30 Dec 2019 14:07:22	U	SMPL
HighStd - 1	HighStd	40271	9.7952 ✓	-	7470A	PPB	30 Dec 2019 14:09:28	U	SMPL
ChkStd - 1	ChkStd	28312	6.8194 ✓	-	7470A	PPB	30 Dec 2019 14:11:35	U	SMPL
PB125782BL - 1	PBW	235	-0.1671 ✓	-	7470A	PPB	30 Dec 2019 14:13:41	U	SMPL
PB125782BS - 1	LCSW	18056	4.2674 ✓	-	7470A	PPB	30 Dec 2019 14:15:48	U	SMPL
K6113-15 - 1	OR-03-122619	852	-0.0135 ✓	-	7470A	PPB	30 Dec 2019 14:17:55	U	SMPL
K6113-15DUP - 1	OR-03-122619DUP	977	0.0176 ✓	-	7470A	PPB	30 Dec 2019 14:20:03	U	SMPL
K6113-15MS - 1	OR-03-122619MS	16769	3.9471 ✓	-	7470A	PPB	30 Dec 2019 14:26:30	U	SMPL
K6113-15MSD - 1	OR-03-122619MSD	15760	3.6961 ✓	-	7470A	PPB	30 Dec 2019 14:30:08	U	SMPL
K6117-12 - 1	TR-05-122619	927	0.0051 ✓	-	7470A	PPB	30 Dec 2019 14:32:14	U	SMPL
CCV84 - 1	CCV84	20456	4.8646 ✓	-	7470A	PPB	30 Dec 2019 14:38:30	U	SMPL
CCB84 - 1	CCB84	127	-0.1940 ✓	-	7470A	PPB	30 Dec 2019 14:40:36	U	SMPL
K6441-08 - 1	K321-WC-01	848	-0.0145 ✓	-	7470A	PPB	30 Dec 2019 14:42:43	U	SMPL
K6441-09 - 1	K321-WC-02	930	0.0059 ✓	-	7470A	PPB	30 Dec 2019 14:44:49	U	SMPL
K6442-07 - 1	K615-WC-1	816	-0.0225 ✓	-	7470A	PPB	30 Dec 2019 14:46:55	U	SMPL
PB125783BL - 1	PBW	630	-0.0688 ✓	-	7470A	PPB	30 Dec 2019 14:49:02	U	SMPL
PB125783BS - 1	LCSW	14857	3.4714 ✓	-	7470A	PPB	30 Dec 2019 14:54:00	U	SMPL
K6114-02 - 1	EO-01-122719	937	0.0076 ✓	-	7470A	PPB	30 Dec 2019 14:56:07	U	SMPL
K6114-02DUP - 1	EO-01-122719DUP	1054	0.0367 ✓	-	7470A	PPB	30 Dec 2019 14:58:14	U	SMPL
K6114-02MS - 1	EO-01-122719MS	17551	4.1417 ✓	-	7470A	PPB	30 Dec 2019 15:00:22	U	SMPL
K6114-02MSD - 1	EO-01-122719MSD	17158	4.0439 ✓	-	7470A	PPB	30 Dec 2019 15:08:18	U	SMPL
K6455-07 - 1	K435-WC-1	1058	0.0377 ✓	-	7470A	PPB	30 Dec 2019 15:10:24	U	SMPL
CCV85 - 1	CCV85	22286	5.3200 ✓	-	7470A	PPB	30 Dec 2019 15:12:31	U	SMPL
CCB85 - 1	CCB85	120	-0.1957 ✓	-	7470A	PPB	30 Dec 2019 15:14:37	U	SMPL
K6456-02 - 1	SB-03	1015	0.0270 ✓	-	7470A	PPB	30 Dec 2019 15:16:43	U	SMPL
K6456-04 - 1	SB-04	1070	0.0407 ✓	-	7470A	PPB	30 Dec 2019 15:18:50	U	SMPL
PB125784BL - 1	PBW	560	-0.0862 ✓	-	7470A	PPB	30 Dec 2019 15:20:56	U	SMPL
PB125784BS - 1	LCSW	18869	4.4697 ✓	-	7470A	PPB	30 Dec 2019 15:23:02	U	SMPL
K6431-15 - 1	GW-1	423	-0.1203 ✓	-	7470A	PPB	30 Dec 2019 15:25:10	U	SMPL
K6431-17 - 1	FIELD-BLANK	249	-0.1636 ✓	-	7470A	PPB	30 Dec 2019 15:27:17	U	SMPL
K6449-01 - 1	LW-04	287	-0.1541 ✓	-	7470A	PPB	30 Dec 2019 15:29:24	U	SMPL
K6449-02 - 1	LW-03	296	-0.1519 ✓	-	7470A	PPB	30 Dec 2019 15:31:32	U	SMPL
K6449-03 - 1	LW-05	318	-0.1464 ✓	-	7470A	PPB	30 Dec 2019 15:33:40	U	SMPL
K6449-04 - 1	OW-4	299	-0.1512 ✓	-	7470A	PPB	30 Dec 2019 15:35:46	U	SMPL
CCV86 - 1	CCV86	22119	5.2784 ✓	-	7470A	PPB	30 Dec 2019 15:40:49	U	SMPL
CCB86 - 1	CCB86	414	-0.1225 ✓	-	7470A	PPB	30 Dec 2019 15:42:56	U	SMPL
K6449-05 - 1	OW-5	311	-0.1482 ✓	-	7470A	PPB	30 Dec 2019 15:45:02	U	SMPL
K6449-05DUP - 1	OW-5DUP	316	-0.1469 ✓	-	7470A	PPB	30 Dec 2019 15:47:09	U	SMPL
K6449-06 - 1	K6449-05MS	16462	3.8708 ✓	-	7470A	PPB	30 Dec 2019 15:53:29	U	SMPL
K6449-07 - 1	K6449-05MSD	15331	3.5893 ✓	-	7470A	PPB	30 Dec 2019 15:55:35	U	SMPL
K6449-08 - 1	OW-2	281	-0.1556 ✓	-	7470A	PPB	30 Dec 2019 15:57:41	U	SMPL
K6449-09 - 1	OW-1	346	-0.1395 ✓	-	7470A	PPB	30 Dec 2019 15:59:48	U	SMPL
K6449-10 - 1	FIELD-BLANK	287	-0.1541 ✓	-	7470A	PPB	30 Dec 2019 16:01:56	U	SMPL
K6449-11 - 1	DUPLICATE	302	-0.1504 ✓	-	7470A	PPB	30 Dec 2019 16:04:03	U	SMPL
K6113-15LX5 - 1		207	-0.1740 ✓	-	7470A	PPB	30 Dec 2019 16:06:12	U	SMPL
K6114-02LX5 - 1		210	-0.1733 ✓	-	7470A	PPB	30 Dec 2019 16:08:18	U	SMPL
CCV87 - 1	CCV87	22514	5.3767 ✓	-	7470A	PPB	30 Dec 2019 16:10:25	U	SMPL
CCB87 - 1	CCB87	352	-0.1380 ✓	-	7470A	PPB	30 Dec 2019 16:12:32	U	SMPL
K6449-05LX5 - 1		72	-0.2076 ✓	-	7470A	PPB	30 Dec 2019 16:14:39	U	SMPL
PB125782TB - 1	PB125746TB	15	-0.2218 ✓	-	7470A	PPB	30 Dec 2019 16:16:46	U	SMPL
PB125783TB - 1	PB125775TB	19	-0.2208 ✓	-	7470A	PPB	30 Dec 2019 16:18:53	U	SMPL
CCV88 - 1	CCV88	20257	4.8151 ✓	-	7470A	PPB	30 Dec 2019 16:21:00	U	SMPL
CCB88 - 1	CCB88	66	-0.2091 ✓	-	7470A	PPB	30 Dec 2019 16:23:07	U	SMPL