



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB106967

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/30/2019	14:30
2	CAL2	1	Water	NA	NA	20.1	7.01	12/30/2019	14:31
3	CAL3	1	Water	NA	NA	20.1	10.03	12/30/2019	14:32
4	ICV	1	Water	NA	NA	20.2	6.99	12/30/2019	14:33
5	CCV1	1	Water	NA	NA	20.1	2.01	12/30/2019	14:35
6	K6469-01	1	Solid	20.03	20	22.1	8.20	12/30/2019	14:37
7	K6469-01DUP	1	Solid	20.04	20	22.2	8.21	12/30/2019	14:38
8	CCV2	1	Water	NA	NA	20.1	12.02	12/30/2019	14:40

106967

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K6469 WorkList ID : 134643 Department : Wet-Chemistry Date : 12-30-2019 14:23:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/30/2019	Solid	K6469-01	pH	Cool 4 deg C	ENTA05	F51	174-NC-TOPSOIL-001	12/30/2019	9045D

Date/Time 12/30/19 14:25
Received by: JA
Relinquished by: CB

Date/Time 12/30/19 15:30
Received by: CB
Relinquished by: JA