



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB106974

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	12/30/2019	14:20
2	K6434-03		1	8	100.00	>212.0	>212.0	12/30/2019	14:45
3	K6434-03DUP		1	8	100.00	>212.0	>212.0	12/30/2019	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB106974

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp123019 WorkList ID : 134641 Department : Wet-Chemistry Date : 12-30-2019 13:52:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2019	Water	K6434-03	Flash Point	Cool 4 deg C	DVIR01	E41	GW-1-DISCHARGE	12/23/2019	1010A

Date/Time 12/30/19 13:55
Received by: KS
Relinquished by: JP

Date/Time 12/30/19 15:15
Received by: JP
Relinquished by: KS