



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106976

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2615
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/30/2019	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	12/30/2019	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	12/30/2019	17:12
4	ICV	1	Water	NA	NA	20.2	7.01	12/30/2019	17:13
5	CCV1	1	Water	NA	NA	20.1	2.01	12/30/2019	17:15
6	K6472-07	1	Solid	20.02	20	20.8	8.35	12/30/2019	17:16
7	K6472-07DUP	1	Solid	20.03	20	20.9	8.36	12/30/2019	17:17
8	K6474-02	1	Solid	20.04	20	20.8	8.20	12/30/2019	17:19
9	K6474-04	1	Solid	20.02	20	21.0	6.86	12/30/2019	17:20
10	CCV2	1	Water	NA	NA	20.1	12.02	12/30/2019	17:22

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K6474 WorkList ID : 134655 Department : Wet-Chemistry Date : 12-30-2019 16:02:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/30/2019	Solid	K6472-07	Corrosivity	Cool 4 deg C	ATCE02	H21	K309-WC-01	12/30/2019	9045D
12/30/2019	Solid	K6474-02	Corrosivity	Cool 4 deg C	PSEG01	H21	SB-01	12/30/2019	9045D
12/30/2019	Solid	K6474-04	Corrosivity	Cool 4 deg C	PSEG01	H21	SB-02	12/30/2019	9045D

Date/Time 12-30-19 16:10
 Received by: JS
 Relinquished by: red

Date/Time 12-30-19 17:13
 Received by: JS
 Relinquished by: JS