



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB106992

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
ZOBELLS SOLUTION ORP STD 500ML	W2615
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	12/31/2019	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	12/31/2019	16:02
3	CAL3	1	Water	NA	NA	20.1	10.03	12/31/2019	16:03
4	ICV	1	Water	NA	NA	20.2	6.99	12/31/2019	16:05
5	CCV1	1	Water	NA	NA	20.1	2.01	12/31/2019	16:06
6	K6113-20	1	Solid	20.03	20	21.1	7.82	12/31/2019	16:07
7	K6113-20DUP	1	Solid	20.04	20	21.2	7.83	12/31/2019	16:08
8	K6479-05	1	Solid	20.02	20	22.6	8.67	12/31/2019	16:10
9	CCV2	1	Water	NA	NA	20.1	12.02	12/31/2019	16:11

✓ 106992

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K6479 WorkList ID : 134690 Department : Wet-Chemistry Date : 12-31-2019 15:16:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2019	Solid	K6113-20	Corrosivity	Cool 4 deg C	PSEG05		OR-02-123119	12/31/2019	9045D
12/30/2019	Solid	K6479-05	Corrosivity	Cool 4 deg C	ATCE02	F31	Q091-WC-01	12/30/2019	9045D

Date/Time 12-31-19 15:20
Received by: JJ
Relinquished by: cd

Date/Time 12-31-19 16:50
Received by: cd
Relinquished by: JR