



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107001

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	01/02/2020	09:15
2	CAL2	1	Water	NA	NA	20.1	7.01	01/02/2020	09:16
3	CAL3	1	Water	NA	NA	20.1	10.04	01/02/2020	09:17
4	ICV	1	Water	NA	NA	20.2	7.00	01/02/2020	09:19
5	CCV1	1	Water	NA	NA	20.2	2.01	01/02/2020	09:20
6	K6484-01	1	Water	NA	NA	14.3	5.21	01/02/2020	09:22
7	K6484-03	1	Water	NA	NA	13.0	5.37	01/02/2020	09:24
8	K6484-05	1	Water	NA	NA	13.5	4.63	01/02/2020	09:25
9	K6484-07	1	Water	NA	NA	13.3	7.46	01/02/2020	09:26
10	K6484-08	1	Water	NA	NA	13.2	6.56	01/02/2020	09:30
11	K6484-09	1	Water	NA	NA	13.9	6.51	01/02/2020	09:32
12	K6484-10	1	Water	NA	NA	15.2	6.49	01/02/2020	09:35
13	K6484-11	1	Water	NA	NA	15.4	5.29	01/02/2020	09:36
14	K6484-12	1	Water	NA	NA	15.0	6.47	01/02/2020	09:37
15	K6484-13	1	Water	NA	NA	15.1	6.30	01/02/2020	09:38
16	CCV2	1	Water	NA	NA	20.2	12.01	01/02/2020	09:39
17	K6484-14	1	Water	NA	NA	14.8	5.45	01/02/2020	09:40
18	K6484-15	1	Water	NA	NA	15.8	5.53	01/02/2020	09:41
19	K6484-15DUP	1	Water	NA	NA	15.9	5.54	01/02/2020	09:42
20	CCV3	1	Water	NA	NA	20.1	2.02	01/02/2020	09:45

WORKLIST(Hardcopy Internal Chain)

107001

WorkList Name : PH W K6484

WorkList ID : 134717

Department : Wet-Chemistry

Date : 01-02-2020 09:06:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2019	Water	K6484-01 ✓	pH	Cool 4 deg C	PSEG01	H33	40121	12/31/2019	9040C
12/31/2019	Water	K6484-03 ✓	pH	Cool 4 deg C	PSEG01	H33	30627	12/31/2019	9040C
12/31/2019	Water	K6484-05 ✓	pH	Cool 4 deg C	PSEG01	H33	30628	12/31/2019	9040C
12/31/2019	Water	K6484-07 ✓	pH	Cool 4 deg C	PSEG01	H33	40330	12/31/2019	9040C
12/31/2019	Water	K6484-08 ✓	pH	Cool 4 deg C	PSEG01	H33	40331	12/31/2019	9040C
12/31/2019	Water	K6484-09 ✓	pH	Cool 4 deg C	PSEG01	H33	40332	12/31/2019	9040C
12/31/2019	Water	K6484-10 ✓	pH	Cool 4 deg C	PSEG01	H33	40333	12/31/2019	9040C
12/31/2019	Water	K6484-11 ✓	pH	Cool 4 deg C	PSEG01	H33	40334	12/31/2019	9040C
12/31/2019	Water	K6484-12 ✓	pH	Cool 4 deg C	PSEG01	H33	40335	12/31/2019	9040C
12/31/2019	Water	K6484-13 ✓	pH	Cool 4 deg C	PSEG01	H33	40336	12/31/2019	9040C
12/31/2019	Water	K6484-14 ✓	pH	Cool 4 deg C	PSEG01	H33	40337	12/31/2019	9040C
12/31/2019	Water	K6484-15 ✓	pH	Cool 4 deg C	PSEG01	H33	40338	12/31/2019	9040C

Date/Time 01.02.20 9:10

Received by: JG

Relinquished by: AM

Date/Time 01.02.20 10:40

Received by: AM

Relinquished by: JG