



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107003

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/02/2020	09:35
2	K6484-01		1	9	100.00	>212.0	>212.0	01/02/2020	09:55
3	K6484-01DUP		1	9	100.00	>212.0	>212.0	01/02/2020	10:25
4	K6484-03		1	8	100.00	>212.0	>212.0	01/02/2020	10:50
5	K6484-05		1	8	100.00	>212.0	>212.0	01/02/2020	11:20
6	K6484-40		1	9	100.00	>212.0	>212.0	01/02/2020	11:45
7	K6484-42		1	7	100.00	>212.0	>212.0	01/02/2020	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 107003

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp1220 WorkList ID : 134716 Department : Wet-Chemistry Date : 01-02-2020 09:06:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/06/2020	Water	K6484-01	F	Cool 4 deg C	PSEG01	H33	40121	12/31/2019	1010A
01/06/2020	Water	K6484-03	F	Cool 4 deg C	PSEG01	H33	30627	12/31/2019	1010A
01/06/2020	Water	K6484-05	F	Cool 4 deg C	PSEG01	H33	30628	12/31/2019	1010A
01/06/2020	Water	K6484-40	D	Cool 4 deg C	PSEG01	H33	40330-40334	12/31/2019	1010A
01/06/2020	Water	K6484-42	D	Cool 4 deg C	PSEG01	H33	40335-40338	12/31/2019	1010A

Date/Time 01/02/2020 09:10
 Received by: KS
 Relinquished by: KS

Date/Time 01/02/2020 12:15
 Received by: KS
 Relinquished by: KS