


LB107007
INSTRUMENT ID (N)

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	-104	0.000	-104				
0.2	0.200	0.201	0.001	943	0.0 %	943				
2.5	2.500	2.547	0.047	11762	0.0 %	11762				
5.0	5.000	4.905	-0.095	22636	0.0 %	22636				
7.5	7.500	7.671	0.171	35393	0.0 %	35393				
10.0	10.000	9.903	-0.097	45689	0.0 %	45689				

LB107007 INSTRUMENT ID CV1

Method: 7470A		Operator: Admin		Date of Analysis: 02 Jan 2020 13:53:10							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1	50.0	-104	-	0.0000	7470A	PPB	02 Jan 2020 14:00:06	S	Std		
0.2 - 1	50.2	943	-	0.2000	7470A	PPB	02 Jan 2020 14:02:14	S	Std		
2.5 - 1	52.5	11762	-	2.5000	7470A	PPB	02 Jan 2020 14:10:41	S	Std		
5.0 - 1	55.0	22636	-	5.0000	7470A	PPB	02 Jan 2020 14:12:48	S	Std		
7.5 - 1	57.5	35393	-	7.5000	7470A	PPB	02 Jan 2020 14:14:55	S	Std		
10.0 - 1	510.0	45689	-	10.0000	7470A	PPB	02 Jan 2020 14:17:01	S	Std		
ICV47 - 1	ICV47	16940	3.6695	-	7470A	PPB	02 Jan 2020 14:19:24	U	SMPL		
ICB47 - 1	ICB47	5	-0.0024	-	7470A	PPB	02 Jan 2020 14:21:29	U	SMPL		
CCV01 - 1	CCV01	22483	4.8714	-	7470A	PPB	02 Jan 2020 14:23:35	U	SMPL		
CCB01 - 1	CCB01	-1	-0.0037	-	7470A	PPB	02 Jan 2020 14:25:41	U	SMPL		
CRA - 1	CRA	1056	0.2255	-	7470A	PPB	02 Jan 2020 14:27:46	U	SMPL		
HighStd - 1	HighStd	46169	10.0071	-	7470A	PPB	02 Jan 2020 14:29:52	U	SMPL		
ChkStd - 1	ChkStd	32479	7.0388	-	7470A	PPB	02 Jan 2020 14:31:58	U	SMPL		
PB125850BL - 1	PBW	-31	-0.0102	-	7470A	PPB	02 Jan 2020 14:34:05	U	SMPL		
PB125850BS - 1	LCSW	17725	3.8397	-	7470A	PPB	02 Jan 2020 14:36:11	U	SMPL		
K6460-04 - 1	GW-6	37	0.0046	-	7470A	PPB	02 Jan 2020 14:38:18	U	SMPL		
K6460-05 - 1	GW-7	41	0.0054	-	7470A	PPB	02 Jan 2020 14:40:25	U	SMPL		
K6461-03 - 1	GW-4	20	0.0009	-	7470A	PPB	02 Jan 2020 14:42:33	U	SMPL		
K6464-10 - 1	GW-1	20	0.0009	-	7470A	PPB	02 Jan 2020 14:44:41	U	SMPL		
K6464-10DUP - 1	GW-1DUP	33	0.0037	-	7470A	PPB	02 Jan 2020 14:46:46	U	SMPL		
CCV02 - 1	CCV02	22702	4.9189	-	7470A	PPB	02 Jan 2020 14:48:52	U	SMPL		
CCB02 - 1	CCB02	9	-0.0015	-	7470A	PPB	02 Jan 2020 14:50:58	U	SMPL		
K6464-11 - 1	K6464-03MS	19581	4.2422	-	7470A	PPB	02 Jan 2020 14:53:04	U	SMPL		
K6464-12 - 1	K6464-03MSD	19282	4.1773	-	7470A	PPB	02 Jan 2020 14:55:10	U	SMPL		
K6464-13 - 1	GW-2	8	-0.0017	-	7470A	PPB	02 Jan 2020 14:57:15	U	SMPL		
K6464-14 - 1	GW-DUPLICATE	19	0.0007	-	7470A	PPB	02 Jan 2020 14:59:21	U	SMPL		
PB125851BL - 1	PBW	-6	-0.0048	-	7470A	PPB	02 Jan 2020 15:01:28	U	SMPL		
PB125851BS - 1	LCSW	18780	4.0685	-	7470A	PPB	02 Jan 2020 15:03:35	U	SMPL		
K6460-02 - 1	GW-6	158	0.0308	-	7470A	PPB	02 Jan 2020 15:05:42	U	SMPL		
K6460-03 - 1	GW-7	195	0.0388	-	7470A	PPB	02 Jan 2020 15:07:49	U	SMPL		
K6461-02 - 1	GW-4	144	0.0278	-	7470A	PPB	02 Jan 2020 15:09:57	U	SMPL		
K6464-03 - 1	GW-1	361	0.0748	-	7470A	PPB	02 Jan 2020 15:12:03	U	SMPL		
CCV03 - 1	CCV03	22940	4.9705	-	7470A	PPB	02 Jan 2020 15:14:09	U	SMPL		
CCB03 - 1	CCB03	-2	-0.0039	-	7470A	PPB	02 Jan 2020 15:16:15	U	SMPL		
K6464-03DUP - 1	GW-1DUP	353	0.0731	-	7470A	PPB	02 Jan 2020 15:18:21	U	SMPL		
K6464-04 - 1	K6464-03MS	19468	4.2177	-	7470A	PPB	02 Jan 2020 15:20:27	U	SMPL		
K6464-05 - 1	K6464-03MSD	19431	4.2096	-	7470A	PPB	02 Jan 2020 15:22:34	U	SMPL		
K6464-06 - 1	GW-2	217	0.0436	-	7470A	PPB	02 Jan 2020 15:24:40	U	SMPL		
K6464-07 - 1	GW-DUPLICATE	247	0.0501	-	7470A	PPB	02 Jan 2020 15:26:47	U	SMPL		
K6464-08 - 1	FIELD-BLANK	155	0.0302	-	7470A	PPB	02 Jan 2020 15:28:54	U	SMPL		
K6482-11 - 1	FB	168	0.0330	-	7470A	PPB	02 Jan 2020 15:31:00	U	SMPL		
K6484-01 - 1	40121	4075	0.8801	-	7470A	PPB	02 Jan 2020 15:33:08	U	SMPL		
K6484-03 - 1	30627	1971	0.4239	-	7470A	PPB	02 Jan 2020 15:35:16	U	SMPL		
K6484-05 - 1	30628	250	0.0508	-	7470A	PPB	02 Jan 2020 15:37:25	U	SMPL		
CCV04 - 1	CCV04	22703	4.9191	-	7470A	PPB	02 Jan 2020 15:39:34	U	SMPL		
CCB04 - 1	CCB04	65	0.0106	-	7470A	PPB	02 Jan 2020 15:41:41	U	SMPL		
K6484-40 - 1	40330-40334	215	0.0432	-	7470A	PPB	02 Jan 2020 15:43:48	U	SMPL		
K6484-42 - 1	40335-40338	175	0.0345	-	7470A	PPB	02 Jan 2020 15:45:54	U	SMPL		
PB125862BL - 1	PBW	-21	-0.0080	-	7470A	PPB	02 Jan 2020 15:48:00	U	SMPL		
PB125862BS - 1	LCSW	18607	4.0310	-	7470A	PPB	02 Jan 2020 15:50:06	U	SMPL		
K6113-20 - 1	OR-02-123119	56	0.0087	-	7470A	PPB	02 Jan 2020 15:52:13	U	SMPL		
K6113-20DUP - 1	OR-02-123119DUP	59	0.0093	-	7470A	PPB	02 Jan 2020 15:54:20	U	SMPL		
K6113-20MS - 1	OR-02-123119MS	19322	4.1860	-	7470A	PPB	02 Jan 2020 15:56:28	U	SMPL		
K6113-20MSD - 1	OR-02-123119MSD	19527	4.2305	-	7470A	PPB	02 Jan 2020 15:58:35	U	SMPL		
K6479-05 - 1	Q091-WC-01	52	0.0078	-	7470A	PPB	02 Jan 2020 16:00:43	U	SMPL		
PB125863BL - 1	PBW	-17	-0.0071	-	7470A	PPB	02 Jan 2020 16:02:50	U	SMPL		
CCV05 - 1	CCV05	23079	5.0006	-	7470A	PPB	02 Jan 2020 16:04:56	U	SMPL		
CCB05 - 1	CCB05	108	0.0200	-	7470A	PPB	02 Jan 2020 16:07:02	U	SMPL		
PB125863BS - 1	LCSW	18810	4.0750	-	7470A	PPB	02 Jan 2020 16:09:08	U	SMPL		
K6472-07 - 1	K309-WC-01	51	0.0076	-	7470A	PPB	02 Jan 2020 16:11:14	U	SMPL		
K6472-07DUP - 1	K309-WC-01DUP	81	0.0141	-	7470A	PPB	02 Jan 2020 16:13:20	U	SMPL		
K6472-07MS - 1	K309-WC-01MS	17750	3.8452	-	7470A	PPB	02 Jan 2020 16:15:27	U	SMPL		
K6472-07MSD - 1	K309-WC-01MSD	17758	3.8469	-	7470A	PPB	02 Jan 2020 16:17:33	U	SMPL		
K6464-10LX5 - 1		-33	-0.0106	-	7470A	PPB	02 Jan 2020 16:19:40	U	SMPL		
K6464-03LX5 - 1		41	0.0054	-	7470A	PPB	02 Jan 2020 16:21:46	U	SMPL		
K6113-20LX5 - 1		-5	-0.0045	-	7470A	PPB	02 Jan 2020 16:23:54	U	SMPL		
K6472-07LX5 - 1		-2	-0.0039	-	7470A	PPB	02 Jan 2020 16:26:01	U	SMPL		
PB125862TB - 1	PB125804TB	20	0.0009	-	7470A	PPB	02 Jan 2020 16:28:08	U	SMPL		
CCV06 - 1	CCV06	22875	4.9564	-	7470A	PPB	02 Jan 2020 16:30:15	U	SMPL		
CCB06 - 1	CCB06	120	0.0226	-	7470A	PPB	02 Jan 2020 16:32:22	U	SMPL		
PB125863TB - 1	PB125836TB	23	0.0015	-	7470A	PPB	02 Jan 2020 16:34:29	U	SMPL		
CCV07 - 1	CCV07	22904	4.9627	-	7470A	PPB	02 Jan 2020 16:36:35	U	SMPL		
CCB07 - 1	CCB07	-2	-0.0039	-	7470A	PPB	02 Jan 2020 16:38:42	U	SMPL		