



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107013

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/02/2020	17:20
2	CAL2	1	Water	NA	NA	20.1	7.01	01/02/2020	17:21
3	CAL3	1	Water	NA	NA	20.1	10.03	01/02/2020	17:22
4	ICV	1	Water	NA	NA	20.2	7.00	01/02/2020	17:24
5	CCV1	1	Water	NA	NA	20.2	2.01	01/02/2020	17:25
6	L1008-02	1	Solid	20.02	20	21.1	7.88	01/02/2020	17:26
7	L1009-02	1	Solid	20.03	20	20.6	8.40	01/02/2020	17:30
8	L1010-02	1	Solid	20.04	20	20.6	7.80	01/02/2020	17:32
9	L1010-02DUP	1	Solid	20.03	20	20.7	7.81	01/02/2020	17:33
10	CCV2	1	Water	NA	NA	20.2	12.02	01/02/2020	17:35

WORKLIST(Hardcopy Internal Chain)

VB107013

WorkList Name : CORROSIONITY L1008

WorkList ID : 134731

Department : Wet-Chemistry

Date : 01-02-2020 17:03:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/02/2020	Solid	L1008-02	Corrosivity	Cool 4 deg C	PSEG05	F21	CL-02-010220	01/02/2020	9045D
01/02/2020	Solid	L1009-02	Corrosivity	Cool 4 deg C	PSEG05	F22	NB-08-010220	01/02/2020	9045D
01/02/2020	Solid	L1010-02	Corrosivity	Cool 4 deg C	PSEG05	F22	OK-01-010220	01/02/2020	9045D

Date/Time 01.02.2020 17:10
 Received by: JM
 Relinquished by: AM

Date/Time 01.02.2020 18:00
 Received by: AM
 Relinquished by: JM