



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107017

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/03/2020	12:40
2	CAL2	1	Water	NA	NA	20.1	7.00	01/03/2020	12:41
3	CAL3	1	Water	NA	NA	20.1	10.03	01/03/2020	12:42
4	ICV	1	Water	NA	NA	20.1	7.00	01/03/2020	12:44
5	CCV1	1	Water	NA	NA	20.1	2.01	01/03/2020	12:45
6	K6484-52	1	Solid	20.03	20	22.1	5.42	01/03/2020	12:46
7	L1028-01	1	Solid	20.04	20	20.8	8.96	01/03/2020	12:48
8	L1028-01DUP	1	Solid	20.03	20	20.9	8.97	01/03/2020	12:49
9	CCV2	1	Water	NA	NA	20.1	12.02	01/03/2020	12:50

VB107017

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1028 WorkList ID : 134761 Department : Wet-Chemistry Date : 01-03-2020 12:23:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/31/2019	Solid	K6484-52	pH	Cool 4 deg C	PSEG01	H33	30544	12/31/2019	9045D
01/03/2020	Solid	L1028-01	pH	Cool 4 deg C	BROO04	F11	NORTH-BERGEN	01/03/2020	9045D

Date/Time 01-03-2020 12:23:45
Received by: JV
Relinquished by: JY

Date/Time 01-03-2020 13:44:00
Received by: JY
Relinquished by: JY