



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107032

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/06/2020	10:30
2	CAL2	1	Water	NA	NA	20.1	6.98	01/06/2020	10:31
3	CAL3	1	Water	NA	NA	20.1	10.03	01/06/2020	10:33
4	ICV	1	Water	NA	NA	20.1	7.00	01/06/2020	10:35
5	CCV1	1	Water	NA	NA	20.1	2.01	01/06/2020	10:36
6	L1033-01	1	Water	NA	NA	18.6	4.69	01/06/2020	10:37
7	L1033-01DUP	1	Water	NA	NA	18.7	4.70	01/06/2020	10:38
8	L1033-02	1	Water	NA	NA	19.1	4.93	01/06/2020	10:39
9	CCV2	1	Water	NA	NA	20.1	12.02	01/06/2020	10:40

2-13 107032

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 11033 WorkList ID : 134776 Department : Wet-Chemistry Date : 01-06-2020 10:21:37

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/02/2020	Water	L1033-01 -17	pH	Cool 4 deg C	STER16	F41	MAIN-CELL	01/02/2020	SM 4500-H B
01/02/2020	Water	L1033-02 -17	pH	Cool 4 deg C	STER16	F41	EX-CELL	01/02/2020	SM 4500-H B

Date/Time 01.06.2020 10:25
Received by: Jo
Relinquished by: Jo

Date/Time 01.06.2020 11:20
Received by: Jo
Relinquished by: Jo