



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107055

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/06/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/06/2020	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	01/06/2020	17:03
4	ICV	1	Water	NA	NA	20.1	6.99	01/06/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	01/06/2020	17:06
6	L1036-01	1	Solid	20.02	20	20.1	11.02	01/06/2020	17:08
7	L1036-02	1	Solid	20.03	20	20.3	11.61	01/06/2020	17:10
8	L1036-02DUP	1	Solid	20.04	20	20.3	11.62	01/06/2020	17:11
9	CCV2	1	Water	NA	NA	20.1	12.02	01/06/2020	17:12

285107055

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1036 WorkList ID : 134812 Department : Wet-Chemistry Date : 01-06-2020 15:31:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/06/2020	Solid	L1036-01	pH	Cool 4 deg C	PSEG01	N11	VB-16182	01/06/2020	9045D
01/06/2020	Solid	L1036-02	pH	Cool 4 deg C	PSEG01	N11	CHRT-25062	01/06/2020	9045D

Date/Time 01-06-2020 15:40
Received by: Ju
Relinquished by: ef

Date/Time 01-06-2020 17:30
Received by: _____
Relinquished by: Ju