


LB107058
INSTRUMENT ID CUI

ISM02.4



Linear

A= 0.0000e+000

B= 1.9450e-004 *slope*

C= 1.9191e-002 *y-intercept*

Rho= 0.9992584

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.027	0.027	40	0.000	40					
0.05	0.050										
0.20	0.200	0.230	0.030	1085	0.0 %	1085					15
2.50	2.500	2.508	0.008	12797	0.0 %	12797					0
5.00	5.000	5.044	0.044	25835	0.0 %	25835					0
7.50	7.500	7.209	-0.291	36966	0.0 %	36966					-4
10.0	10.000	10.181	0.181	52247	0.0 %	52247					2

6
01/07/20

LB107058
INSTRUMENT ID:CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0	S0	40 -	0	ISM02.4	PPB	1/7/2020 9:57	Std
0.2	S0.2	1085 -	0.2	ISM02.4	PPB	1/7/2020 9:59	Std
2.5	S2.5	12797 -	2.5	ISM02.4	PPB	1/7/2020 10:01	Std
5	S5.0	25835 -	5	ISM02.4	PPB	1/7/2020 10:04	Std
7.5	S7.5	36966 -	7.5	ISM02.4	PPB	1/7/2020 10:06	Std
10	S10.0	52247 -	10	ISM02.4	PPB	1/7/2020 10:08	Std
ICV053	ICV053	19746 3.8598 -		ISM02.4	PPB	1/7/2020 10:18	SMPL
ICB053	ICB053	-39 0.0116 -		ISM02.4	PPB	1/7/2020 10:20	SMPL
CCV021	CCV021	25660 5.0101 -		ISM02.4	PPB	1/7/2020 10:22	SMPL
CCB021	CCB021	-36 0.0122 -		ISM02.4	PPB	1/7/2020 10:25	SMPL
PB125918E	PBS918	-16 0.0161 -		ISM02.4	PPB	1/7/2020 10:27	SMPL
K6445-19	X2K11	29162 5.6912 -		ISM02.4	PPB	1/7/2020 10:29	SMPL
K6445-20	X2K11D	29074 5.6741 -		ISM02.4	PPB	1/7/2020 10:31	SMPL
K6445-21	X2K11S	42276 8.2419 -		ISM02.4	PPB	1/7/2020 10:34	SMPL
CCV022	CCV022	25625 5.0033 -		ISM02.4	PPB	1/7/2020 10:36	SMPL
CCB022	CCB022	-52 0.0091 -		ISM02.4	PPB	1/7/2020 10:38	SMPL