


LB107059

INSTRUMENT ID CV1

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	-52	0.000	-52				
0.2	0.200	0.231	0.031	1303	0.0 %	1303				
2.5	2.500	2.521	0.021	15830	0.0 %	15830				
5.0	5.000	5.036	0.036	31782	0.0 %	31782				
7.5	7.500	7.216	-0.284	45617	0.0 %	45617				
10.0	10.000	10.177	0.177	64403	0.0 %	64403				

LB107059 INSTRUMENT ID CV1

Method: 7471B		Operator: Admin		Date of Analysis: 07 Jan 2020 11:13:15					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	So.0	-52	-	0.0000	7471B	PPB	07 Jan 2020 11:13:54	S	Std
0.2 - 1	So.2	1303	-	0.2000	7471B	PPB	07 Jan 2020 11:16:11	S	Std
2.5 - 1	S2.5	15830	-	2.5000	7471B	PPB	07 Jan 2020 11:18:28	S	Std
5.0 - 1	S5.0	31782	-	5.0000	7471B	PPB	07 Jan 2020 11:20:45	S	Std
7.5 - 1	S7.5	45617	-	7.5000	7471B	PPB	07 Jan 2020 11:23:02	S	Std
10.0 - 1	S10.0	64403	-	10.0000	7471B	PPB	07 Jan 2020 11:48:03	S	Std
ICV54 - 1	ICV54	22980	3.6482 ✓	-	7471B	PPB	07 Jan 2020 11:52:16	U	SMPL
ICB54 - 1	ICB54	-75	0.0142 ✓	-	7471B	PPB	07 Jan 2020 11:54:31	U	SMPL
CCV23 - 1	CCV23	31416	4.9780 ✓	-	7471B	PPB	07 Jan 2020 11:56:45	U	SMPL
CCB23 - 1	CCB23	-108	0.0090 ✓	-	7471B	PPB	07 Jan 2020 11:59:00	U	SMPL
CRA - 1	CRA	1211	0.2169 ✓	-	7471B	PPB	07 Jan 2020 12:01:16	U	SMPL
HighStd - 1	HighStd	58967	9.3206 ✓	-	7471B	PPB	07 Jan 2020 12:07:51	U	SMPL
ChkStd - 1	ChkStd	43202	6.8357 ✓	-	7471B	PPB	07 Jan 2020 12:10:06	U	SMPL
PB125922BL - 1	PBS	-129	0.0057 ✓	-	7471B	PPB	07 Jan 2020 12:12:22	U	SMPL
PB125922BS - 1	LCSS	24097	3.8243 ✓	-	7471B	PPB	07 Jan 2020 12:14:39	U	SMPL
L1017-01 - 1	TR-04-010620	5443	0.8840 ✓	-	7471B	PPB	07 Jan 2020 12:16:55	U	SMPL
L1032-01 - 1	PAMT-COMP-3-1-2-20	4280	0.7007 ✓	-	7471B	PPB	07 Jan 2020 12:19:11	U	SMPL
L1032-01DUP - 1	PAMT-COMP-3-1-2-20DUP	4198	0.6878 ✓	-	7471B	PPB	07 Jan 2020 12:21:28	U	SMPL
L1032-01MS - 1	PAMT-COMP-3-1-2-20MS	27231	4.3183 ✓	-	7471B	PPB	07 Jan 2020 12:23:45	U	SMPL
L1032-01MSD - 1	PAMT-COMP-3-1-2-20MSD	25772	4.0883 ✓	-	7471B	PPB	07 Jan 2020 12:26:00	U	SMPL
CCV24 - 1	CCV24	31567	5.0018 ✓	-	7471B	PPB	07 Jan 2020 12:28:15	U	SMPL
CCB24 - 1	CCB24	-139	0.0041 ✓	-	7471B	PPB	07 Jan 2020 12:30:30	U	SMPL
L1032-02 - 1	PAMT-COMP-4-1-2-20	6914	1.1159 ✓	-	7471B	PPB	07 Jan 2020 12:32:45	U	SMPL
L1032-03 - 1	PAMT-COMP-5-1-2-20	987	0.1816 ✓	-	7471B	PPB	07 Jan 2020 12:35:00	U	SMPL
L1032-04 - 1	PIER-80N-CORE-1-1-3-20	7805	1.2563 ✓	-	7471B	PPB	07 Jan 2020 12:37:15	U	SMPL
L1032-05 - 1	PIER-80S-CORE-1-1-3-20	6892	1.1124 ✓	-	7471B	PPB	07 Jan 2020 12:39:31	U	SMPL
L1032-06 - 1	PIER-82S-CORE-1-1-3-20	7658	1.2331 ✓	-	7471B	PPB	07 Jan 2020 12:41:47	U	SMPL
L1032-07 - 1	PIER-84N-CORE-1-1-3-20	7674	1.2357 ✓	-	7471B	PPB	07 Jan 2020 12:44:02	U	SMPL
L1032-08 - 1	KIND-CORE-1-1-3-20	9891	1.5851 ✓	-	7471B	PPB	07 Jan 2020 12:46:18	U	SMPL
L1032-09 - 1	DUP	5971	0.9672 ✓	-	7471B	PPB	07 Jan 2020 12:48:35	U	SMPL
L1032-01LX5 - 1		597	0.1202 ✓	-	7471B	PPB	07 Jan 2020 12:50:52	U	SMPL
CCV25 - 1	CCV25	31991	5.0686 ✓	-	7471B	PPB	07 Jan 2020 12:53:08	U	SMPL
CCB25 - 1	CCB25	-133	0.0051 ✓	-	7471B	PPB	07 Jan 2020 12:55:23	U	SMPL