



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107060

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	3.99	01/07/2020	10:10
2	CAL2	1	Water	NA	NA	20.2	7.01	01/07/2020	10:11
3	CAL3	1	Water	NA	NA	20.2	10.03	01/07/2020	10:12
4	ICV	1	Water	NA	NA	20.1	7.00	01/07/2020	10:15
5	CCV1	1	Water	NA	NA	20.1	2.02	01/07/2020	10:16
6	L1041-01	1	Water	NA	NA	15.8	6.72	01/07/2020	10:18
7	L1041-01DUP	1	Water	NA	NA	15.9	6.73	01/07/2020	10:19
8	L1041-04	1	Water	NA	NA	17.0	6.74	01/07/2020	10:20
9	CCV2	1	Water	NA	NA	20.1	12.02	01/07/2020	10:22

WORKLIST(Hardcopy Internal Chain)

13101060

WorkList Name : PH L1041 WorkList ID : 134828 Department : Wet-Chemistry Date : 01-07-2020 10:00:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/06/2020	Water	L1041-01 ✓	pH	Cool 4 deg C	APT101	F11	WGGM-20-WW-TSF-002	01/06/2020	SM 4500-H B
01/06/2020	Water	L1041-04 ✓	pH	Cool 4 deg C	APT101	F11	WGGM-20-WW-TSF-003	01/06/2020	SM 4500-H B

Date/Time 01-07-2020 10:05
 Received by: Jo
 Relinquished by: Av

Date/Time 01-07-2020 11:00
 Received by: AG
 Relinquished by: Jg