

Analysis Method: 9034
 Parameter: Reactive Sulfide
 Run Number: LB107064

ANALYST: Alexander
 SUPERVISOR REVIEW BY: apatel
 Constant: 16000
 Normality1: 0.025
 Normality2: 0.025

Reagent/Standard	Lot/Log #
Sodium thiosulfate titrant 0.025N	WP80247
IODINE SOLUTION .025N 1L	W2508
Starch Solution, 4L	W2379

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB125934BL		1	5.00	50	2.00	0.00	2.00	2.00	0.00	0.00	0.00	01/07/2020	12:10
2	L1017-03		1	5.05	50	2.00	2.00	3.96	1.96	0.04	0.04	3.17	01/07/2020	12:12
3	L1017-03DUP		1	5.03	50	2.00	3.96	5.96	2.00	0.00	0.00	0.00	01/07/2020	12:14
4	L1039-03		1	5.05	50	2.00	5.96	7.84	1.88	0.12	0.12	9.50	01/07/2020	12:16
5	L1039-04		1	5.06	50	2.00	0.00	1.90	1.90	0.10	0.10	7.91	01/07/2020	12:20
6	L1039-07		1	5.09	50	2.00	1.90	3.82	1.92	0.08	0.08	6.29	01/07/2020	12:22
7	L1039-08		1	5.03	50	2.00	3.82	5.72	1.90	0.10	0.10	7.95	01/07/2020	12:24
8	L1039-09		1	5.04	50	2.00	5.72	7.30	1.58	0.42	0.42	33.33	01/07/2020	12:26

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume