



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107071

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/07/2020	17:05
2	CAL2	1	Water	NA	NA	20.2	7.01	01/07/2020	17:06
3	CAL3	1	Water	NA	NA	20.2	10.03	01/07/2020	17:07
4	ICV	1	Water	NA	NA	20.2	6.99	01/07/2020	17:09
5	CCV1	1	Water	NA	NA	20.2	2.02	01/07/2020	17:10
6	L1013-06	1	Solid	20.03	20	20.8	8.45	01/07/2020	17:12
7	L1016-04	1	Solid	20.04	20	20.3	7.71	01/07/2020	17:15
8	L1016-04DUP	1	Solid	20.03	20	20.4	7.72	01/07/2020	17:16
9	CCV2	1	Water	NA	NA	20.1	12.02	01/07/2020	17:20

107071

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K641

WorkList ID : 134834

Department : Wet-Chemistry

Date : 01-07-2020 14:18:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/07/2020	Solid	L1013-06	Corrosivity	Cool 4 deg C	PSEG05	I12	PL-01-010720	01/07/2020	9045D
01/07/2020	Solid	L1016-04	Corrosivity	Cool 4 deg C	PSEG05	I12	SU-02-010720	01/07/2020	9045D

Date/Time 01.07.2020 14:20
 Received by: Jo
 Relinquished by: Le

Date/Time 01.07.2020 14:30
 Received by: Jo
 Relinquished by: Jo