



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107072

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/07/2020	14:30
2	CAL2	1	Water	NA	NA	20.1	7.01	01/07/2020	14:31
3	CAL3	1	Water	NA	NA	20.1	10.03	01/07/2020	14:33
4	ICV	1	Water	NA	NA	20.1	6.99	01/07/2020	14:35
5	CCV1	1	Water	NA	NA	20.2	2.02	01/07/2020	15:15
6	L1049-01	1	Solid	20.02	20	20.1	9.64	01/07/2020	15:16
7	L1049-01DUP	1	Solid	20.03	20	20.2	9.65	01/07/2020	15:17
8	L1049-04	1	Solid	20.05	20	18.4	11.23	01/07/2020	15:20
9	L1049-07	1	Solid	20.03	20	20.1	8.41	01/07/2020	15:22
10	L1049-10	1	Solid	20.04	20	19.8	6.16	01/07/2020	15:24
11	CCV2	1	Water	NA	NA	20.1	12.02	01/07/2020	15:25

13107072

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1049 S WorkList ID : 134837 Department : Wet-Chemistry Date : 01-07-2020 15:04:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/07/2020	Solid	L1049-01	✓ pH	Cool 4 deg C	PSEG01	I12	TP-1	01/07/2020	9045D
01/07/2020	Solid	L1049-04	✓ pH	Cool 4 deg C	PSEG01	I12	TP-2	01/07/2020	9045D
01/07/2020	Solid	L1049-07	✓ pH	Cool 4 deg C	PSEG01	I12	TP-3	01/07/2020	9045D
01/07/2020	Solid	L1049-10	✓ pH	Cool 4 deg C	PSEG01	I12	TP-4	01/07/2020	9045D

Date/Time 01.07.2020 15:10
 Received by: Jø
 Relinquished by: Jø

Date/Time 01.07.2020 17:30
 Received by: J.C
 Relinquished by: Jø