


LB107073

INST RUNMENT ID C01

7470A



Linear

A= 0.0000e+000

B= 2.0821e-004

C= -6.9226e-003

Rho= 0.9999723

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-22	0.000	-22				
0.2	0.200	0.201	0.001	1001	0.0 %	1001				
2.5	2.500	2.502	0.002	12049	0.0 %	12049				
5.0	5.000	4.991	-0.009	24006	0.0 %	24006				
7.5	7.500	7.554	0.054	36315	0.0 %	36315				
10.0	10.000	9.963	-0.037	47884	0.0 %	47884				

LB107073 INSTRUMENT ID CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Jan 2020 07:57:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	-22	-	0.0000	7470A	PPB	08 Jan 2020 08:10:44	S	Std
0.2 - 1	60.2	1001	-	0.2000	7470A	PPB	08 Jan 2020 08:12:51	S	Std
2.5 - 1	52.5	12049	-	2.5000	7470A	PPB	08 Jan 2020 08:14:58	S	Std
5.0 - 1	55.0	24006	-	5.0000	7470A	PPB	08 Jan 2020 08:17:06	S	Std
7.5 - 1	57.5	36315	-	7.5000	7470A	PPB	08 Jan 2020 08:19:13	S	Std
10.0 - 1	510.0	47884	-	10.0000	7470A	PPB	08 Jan 2020 08:21:19	S	Std
ICV55 - 1	ICV55	18383	3.8205 ✓	-	7470A	PPB	08 Jan 2020 08:23:49	U	SMPL
ICB55 - 1	ICB55	-72	-0.0219 ✓	-	7470A	PPB	08 Jan 2020 08:25:55	U	SMPL
CCV26 - 1	CCV26	23637	4.9145 ✓	-	7470A	PPB	08 Jan 2020 08:28:00	U	SMPL
CCB26 - 1	CCB26	-51	-0.0175 ✓	-	7470A	PPB	08 Jan 2020 08:30:06	U	SMPL
CRA - 1	CRA	1027	0.2069 ✓	-	7470A	PPB	08 Jan 2020 08:32:11	U	SMPL
HighStd - 1	HighStd	47922	9.9708 ✓	-	7470A	PPB	08 Jan 2020 08:34:17	U	SMPL
ChkStd - 1	ChkStd	33051	6.8745 ✓	-	7470A	PPB	08 Jan 2020 08:36:23	U	SMPL
PB125938BL - 1	PBW	-49	-0.0171 ✓	-	7470A	PPB	08 Jan 2020 08:38:29	U	SMPL
PB125938BS - 1	LCSW	20371	4.2345 ✓	-	7470A	PPB	08 Jan 2020 08:40:36	U	SMPL
L1017-03 - 1	TR-04-010620	-10	-0.0090 ✓	-	7470A	PPB	08 Jan 2020 08:42:43	U	SMPL
L1017-03DUP - 1	TR-04-010620DUP	11	-0.0046 ✓	-	7470A	PPB	08 Jan 2020 08:44:50	U	SMPL
L1017-03MS - 1	TR-04-010620MS	19319	4.0154 ✓	-	7470A	PPB	08 Jan 2020 08:46:57	U	SMPL
L1017-03MSD - 1	TR-04-010620MSD	19492	4.0514 ✓	-	7470A	PPB	08 Jan 2020 08:49:05	U	SMPL
L1038-01 - 1	01-06-A	4	-0.0061 ✓	-	7470A	PPB	08 Jan 2020 08:51:12	U	SMPL
CCV27 - 1	CCV27	23616	4.9101 ✓	-	7470A	PPB	08 Jan 2020 08:53:17	U	SMPL
CCB27 - 1	CCB27	-15	-0.0100 ✓	-	7470A	PPB	08 Jan 2020 08:55:23	U	SMPL
L1038-02 - 1	01-06-B	20	-0.0028 ✓	-	7470A	PPB	08 Jan 2020 08:57:29	U	SMPL
L1039-03 - 1	301	17	-0.0034 ✓	-	7470A	PPB	08 Jan 2020 08:59:35	U	SMPL
L1039-04 - 1	305	38	0.0010 ✓	-	7470A	PPB	08 Jan 2020 09:01:41	U	SMPL
L1039-07 - 1	225	55	0.0045 ✓	-	7470A	PPB	08 Jan 2020 09:03:47	U	SMPL
L1039-08 - 1	230	57	0.0049 ✓	-	7470A	PPB	08 Jan 2020 09:05:54	U	SMPL
L1017-03LX5 - 1	-	-1	-0.0071 ✓	-	7470A	PPB	08 Jan 2020 09:08:01	U	SMPL
PB125938TB - 1	PB125923TB	0	-0.0069 ✓	-	7470A	PPB	08 Jan 2020 09:10:07	U	SMPL
CCV28 - 1	CCV28	23643	4.9157 ✓	-	7470A	PPB	08 Jan 2020 09:12:15	U	SMPL
CCB28 - 1	CCB28	-9	-0.0088 ✓	-	7470A	PPB	08 Jan 2020 09:14:22	U	SMPL