



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107086

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/08/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.00	01/08/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/08/2020	16:02
4	ICV	1	Water	NA	NA	20.1	6.99	01/08/2020	16:03
5	CCV1	1	Water	NA	NA	20.2	2.01	01/08/2020	16:05
6	L1060-06	1	Solid	20.03	20	20.1	6.60	01/08/2020	16:06
7	L1060-12	1	Solid	20.04	20	19.8	6.77	01/08/2020	16:07
8	L1060-18	1	Solid	20.02	20	19.4	6.73	01/08/2020	16:09
9	L1060-18DUP	1	Solid	20.01	20	19.5	6.74	01/08/2020	16:10
10	CCV2	1	Water	NA	NA	20.1	12.02	01/08/2020	16:13

WORKLIST(Hardcopy Internal Chain)

107086

WorkList Name : CORROSIVITY L1060

WorkList ID : 134878

Department : Wet-Chemistry

Date : 01-08-2020 15:12:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2020	Solid	L1060-06	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-1	01/08/2020	9045D
01/08/2020	Solid	L1060-12	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-2	01/08/2020	9045D
01/08/2020	Solid	L1060-18	Corrosivity	Cool 4 deg C	PSEG01	I13	TP-3	01/08/2020	9045D

Date/Time 01-08-2020 15:20
 Received by: SA CP
 Relinquished by: _____

Date/Time 01-08-2020 17:30
 Received by: CP
 Relinquished by: J2