



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107098

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/09/2020	09:20
2	L1047-02		1	9	100.00	>212.0	>212.0	01/09/2020	09:45
3	L1047-02DUP		1	9	100.00	>212.0	>212.0	01/09/2020	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp1719 WorkList ID : 134829 Department : Wet-Chemistry Date : 01-07-2020 11:02:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/14/2020	Water	L1047-02 C	Flash Point	Cool 4 deg C	PSEG01	I12	FMI109	01/07/2020	1010A

Date/Time 01/09/20 08:45
 Received by: KS
 Relinquished by: DL

Date/Time 01/09/20 10:30
 Received by: DL
 Relinquished by: KS