



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107102

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/09/2020	13:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/09/2020	13:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/09/2020	13:02
4	ICV	1	Water	NA	NA	20.1	6.99	01/09/2020	13:05
5	CCV1	1	Water	NA	NA	20.1	2.01	01/09/2020	13:06
6	L1062-01	1	Solid	20.02	20	21.1	6.41	01/09/2020	13:10
7	L1062-01DUP	1	Solid	20.03	20	21.1	6.42	01/09/2020	13:11
8	L1069-01	1	Solid	20.05	20	20.5	10.51	01/09/2020	13:12
9	L1069-02	1	Solid	20.02	20	20.0	11.41	01/09/2020	13:15
10	CCV2	1	Water	NA	NA	20.2	12.02	01/09/2020	13:20

28107102

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1069 WorkList ID : 134892 Department : Wet-Chemistry Date : 01-09-2020 12:40:57

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/09/2020	Solid	L1062-01	pH	Cool 4 deg C	PSEG01	I21	RT-3848	01/09/2020	9045D
01/09/2020	Solid	L1069-01	pH	Cool 4 deg C	PSEG01	I21	VB16182	01/09/2020	9045D
01/09/2020	Solid	L1069-02	pH	Cool 4 deg C	PSEG01	I21	CHRT2506	01/09/2020	9045D

Date/Time 01.09.2020 12:50
Received by: So
Relinquished by: DK

Date/Time 01.09.2020 13:20
Received by: DK
Relinquished by: Jm