



## Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107117

Slope : 98.4

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2628         |
| buffer solution pH 7 yellow    | W2626         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2569         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2616         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.1             | 4.01        | 01/10/2020 | 13:05     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.1             | 7.00        | 01/10/2020 | 13:06     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.1             | 10.03       | 01/10/2020 | 13:07     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 7.01        | 01/10/2020 | 13:10     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.01        | 01/10/2020 | 13:11     |
| 6   | L1087-01    | 1  | Water  | NA          | NA          | 17.2             | 5.16        | 01/10/2020 | 13:12     |
| 7   | L1087-01DUP | 1  | Water  | NA          | NA          | 17.3             | 5.17        | 01/10/2020 | 13:14     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 01/10/2020 | 13:15     |

# WORKLIST(Hardcopy Internal Chain)

20107117

WorkList Name : PH L1087      WorkList ID : 134941      Department : Wet-Chemistry      Date : 01-10-2020 12:45:14

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method      |
|------------|--------|-------------|------|--------------|----------|------------------|-----------------|--------------|-------------|
| 01/06/2020 | Water  | L1087-01-17 | pH   | Cool 4 deg C | STER11   | I13              | POOL-WATER      | 01/06/2020   | SM 4500-H B |

Date/Time 01-10-2020 12:50  
 Received by: SO  
 Relinquished by: SO

Date/Time 01-10-2020 14:00  
 Received by: SO  
 Relinquished by: SO