

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB107125

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP81082
calibration std. phosphate 0.5 ppm	WP81081
calibration std. phosphate 0.3 ppm	WP81080
calibration std. phosphate 0.1 ppm	WP81079
calibration std. phosphate 0.05 ppm	WP81078
calibration std. 0 ppm	WP81077
phosphate CCV std.	WP81083
5N sulfuric acid	WP79628
Combined reagent	WP81127
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314

Intercept: -0.0005Slope: 0.6491Regression: 0.999823

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/10/2020	12:26
2	CAL2	0.05	1	50	50	0.033	01/10/2020	12:27
3	CAL3	0.10	1	50	50	0.066	01/10/2020	12:28
4	CAL4	0.30	1	50	50	0.186	01/10/2020	12:28
5	CAL5	0.50	1	50	50	0.330	01/10/2020	12:29
6	CAL6	1.00	1	50	50	0.648	01/10/2020	12:30



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB107125

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.323	0.498	01/10/2020	12:30
2	ICB		1	50	50	0.000	0.001	01/10/2020	12:31
3	CCV1	0.50	1	50	50	0.328	0.506	01/10/2020	12:31
4	CCB1		1	50	50	0.000	0.001	01/10/2020	12:32
5	PB126020BL		1	50	50	0.000	0.001	01/10/2020	12:32
6	PB126020BS	0.50	1	50	50	0.349	0.538	01/10/2020	12:33
7	L1035-01		1	50	50	0.785	1.210	01/10/2020	12:34
8	L1035-01DUP		1	50	50	0.746	1.150	01/10/2020	12:34
9	L1035-01MS	0.50	1	50	50	0.893	1.377	01/10/2020	12:36
10	L1035-01MSD	0.50	1	50	50	0.846	1.304	01/10/2020	12:37
11	L1035-01		5	50	50	0.249	0.384	01/10/2020	13:05
12	L1035-01DUP		5	50	50	0.233	0.360	01/10/2020	13:05
13	L1035-01MS	0.50	5	50	50	0.303	0.468	01/10/2020	13:06
14	L1035-01MSD	0.50	5	50	50	0.299	0.461	01/10/2020	13:06
15	CCV2	0.50	1	50	50	0.329	0.508	01/10/2020	13:07
16	CCB2		1	50	50	0.000	0.001	01/10/2020	13:07