


LB107126
INSTRUMENT ID: CV1

7471B

Linear

μ Abs.:

62814



A= 0.0000e+000

B= 1.6147e-004

C= 2.0613e-002

Rho= 0.9986260

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.080	0.080	365	0.000	365				
0.2	0.200	0.334	0.134	1943	0.0 %	1943				
2.5	2.500	2.519	0.019	15471	0.0 %	15471				
5.0	5.000	4.584	-0.416	28265	0.0 %	28265				
7.5	7.500	7.520	0.020	46445	0.0 %	46445				
10.0	10.000	10.163	0.163	62814	0.0 %	62814				

LB107126 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Jan 2020 14:13:56

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	365	-	0.0000	7471B	PPB	10 Jan 2020 14:14:14	S	Std
0.2 - 1	SD-2	1943	-	0.2000	7471B	PPB	10 Jan 2020 14:16:31	S	Std
2.5 - 1	SD-5	15471	-	2.5000	7471B	PPB	10 Jan 2020 14:18:48	S	Std
5.0 - 1	SD-10	28265	-	5.0000	7471B	PPB	10 Jan 2020 14:21:05	S	Std
7.5 - 1	SD-15	46445	-	7.5000	7471B	PPB	10 Jan 2020 14:23:21	S	Std
10.0 - 1	SD-20	62814	-	10.0000	7471B	PPB	10 Jan 2020 14:25:37	S	Std
ICV65 - 1	ICV65	22976	3.7305	-	7471B	PPB	10 Jan 2020 14:32:44	U	SMPL
ICB65 - 1	ICB65	345	0.0763	-	7471B	PPB	10 Jan 2020 14:34:58	U	SMPL
CCV37 - 1	CCV37	31196	5.0577	-	7471B	PPB	10 Jan 2020 14:37:13	U	SMPL
CCB37 - 1	CCB37	314	0.0713	-	7471B	PPB	10 Jan 2020 14:39:29	U	SMPL
CRA - 1	CRA	1289	0.2287	-	7471B	PPB	10 Jan 2020 14:41:45	U	SMPL
HighStd - 1	HighStd	61494	9.9499	-	7471B	PPB	10 Jan 2020 14:44:00	U	SMPL
ChkStd - 1	ChkStd	41372	6.7008	-	7471B	PPB	10 Jan 2020 14:46:15	U	SMPL
PB126013BL - 1	PBS	300	0.0691	-	7471B	PPB	10 Jan 2020 14:48:30	U	SMPL
PB126013BS - 1	LCSS	22334	3.6268	-	7471B	PPB	10 Jan 2020 14:50:46	U	SMPL
L1060-01 - 1	TP-1	6772	1.1141	-	7471B	PPB	10 Jan 2020 14:53:02	U	SMPL
L1060-01DUP - 1	TP-1DUP	6890	1.1331	-	7471B	PPB	10 Jan 2020 14:55:19	U	SMPL
L1060-01MS - 1	TP-1MS	32110	5.2053	-	7471B	PPB	10 Jan 2020 15:06:28	U	SMPL
L1060-01MSD - 1	TP-1MSD	32445	5.2594	-	7471B	PPB	10 Jan 2020 15:08:45	U	SMPL
L1060-07 - 1	TP-2	7167	1.1778	-	7471B	PPB	10 Jan 2020 15:11:00	U	SMPL
CCV38 - 1	CCV38	31436	5.0965	-	7471B	PPB	10 Jan 2020 15:13:15	U	SMPL
CCB38 - 1	CCB38	294	0.0681	-	7471B	PPB	10 Jan 2020 15:15:30	U	SMPL
L1060-13 - 1	TP-3	4680	0.7763	-	7471B	PPB	10 Jan 2020 15:17:46	U	SMPL
L1064-01 - 1	SOIL-A	1053	0.1906	-	7471B	PPB	10 Jan 2020 15:20:01	U	SMPL
L1064-05 - 1	SOIL-B	1220	0.2176	-	7471B	PPB	10 Jan 2020 15:22:16	U	SMPL
L1064-09 - 1	OD	1699	0.2949	-	7471B	PPB	10 Jan 2020 15:24:32	U	SMPL
L1070-01 - 1	RBR-200030	2634	0.4459	-	7471B	PPB	10 Jan 2020 15:26:47	U	SMPL
L1070-04 - 1	RBR-200059	1795	0.3104	-	7471B	PPB	10 Jan 2020 15:29:04	U	SMPL
L1084-01 - 1	AF-1	25427	4.1262	-	7471B	PPB	10 Jan 2020 15:31:20	U	SMPL
L1084-02 - 1	AF-2	24951	4.0494	-	7471B	PPB	10 Jan 2020 15:33:36	U	SMPL
L1084-03 - 1	AF-3	24109	3.9134	-	7471B	PPB	10 Jan 2020 15:35:54	U	SMPL
L1060-01LX5 - 1		1334	0.2360	-	7471B	PPB	10 Jan 2020 15:38:10	U	SMPL
CCV39 - 1	CCV39	31705	5.1399	-	7471B	PPB	10 Jan 2020 15:40:26	U	SMPL
CCB39 - 1	CCB39	315	0.0715	-	7471B	PPB	10 Jan 2020 15:42:42	U	SMPL