

LB107131


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47993



A= 0.0000e+000

B= 2.0552e-004

C= -5.6182e-002

Rho= 0.9982641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.022	0.022	380	0.000	380				
0.2	0.200	0.198	-0.002	1235	0.0 %	1235				
2.5	2.500	2.579	0.079	12821	0.0 %	12821				
5.0	5.000	4.704	-0.296	23163	0.0 %	23163				
7.5	7.500	7.890	0.390	38664	0.0 %	38664				
10.0	10.000	9.807	-0.193	47993	0.0 %	47993				

LB107131 INSTRUMENT ID:CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jan 2020 17:16:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	380	-	0.0000	7470A	PPB	10 Jan 2020 17:16:29	S	Std
0.2 - 1	50.2	1235	-	0.2000	7470A	PPB	10 Jan 2020 17:18:37	S	Std
2.5 - 1	52.5	12821	-	2.5000	7470A	PPB	10 Jan 2020 17:20:44	S	Std
5.0 - 1	55.0	23163	-	5.0000	7470A	PPB	10 Jan 2020 17:22:52	S	Std
7.5 - 1	57.5	38664	-	7.5000	7470A	PPB	10 Jan 2020 17:24:59	S	Std
10.0 - 1	59.0	47993	-	10.0000	7470A	PPB	10 Jan 2020 17:34:28	S	Std
ICV66 - 1	ICV66	19679	3.9882	-	7470A	PPB	10 Jan 2020 17:42:27	U	SMPL
ICB66 - 1	ICB66	303	0.0061	-	7470A	PPB	10 Jan 2020 17:44:32	U	SMPL
CCV40 - 1	CCV40	25851	5.2567	-	7470A	PPB	10 Jan 2020 17:46:37	U	SMPL
CCB40 - 1	CCB40	287	0.0028	-	7470A	PPB	10 Jan 2020 17:48:43	U	SMPL
CRA - 1	CRA	1205	0.1915	-	7470A	PPB	10 Jan 2020 17:50:48	U	SMPL
HighStd - 1	HighStd	50649	10.3532	-	7470A	PPB	10 Jan 2020 17:52:54	U	SMPL
ChkStd - 1	ChkStd	34807	7.0974	-	7470A	PPB	10 Jan 2020 17:58:58	U	SMPL
PB126014BL - 1	PBS	255	-0.0038	-	7470A	PPB	10 Jan 2020 18:01:04	U	SMPL
PB126014BS - 1	LCSS	18216	3.6876	-	7470A	PPB	10 Jan 2020 18:03:10	U	SMPL
L1060-06 - 1	TP-1	145	-0.0264	-	7470A	PPB	10 Jan 2020 18:05:17	U	SMPL
L1060-06DUP - 1	TP-1DUP	157	-0.0239	-	7470A	PPB	10 Jan 2020 18:07:24	U	SMPL
L1060-06MS - 1	TP-1MS	16329	3.2998	-	7470A	PPB	10 Jan 2020 18:09:31	U	SMPL
L1060-06MSD - 1	TP-1MSD	16816	3.3998	-	7470A	PPB	10 Jan 2020 18:11:39	U	SMPL
L1060-12 - 1	TP-2	128	-0.0299	-	7470A	PPB	10 Jan 2020 18:13:45	U	SMPL
CCV41 - 1	CCV41	23694	4.8134	-	7470A	PPB	10 Jan 2020 18:15:51	U	SMPL
CCB41 - 1	CCB41	290	0.0034	-	7470A	PPB	10 Jan 2020 18:17:58	U	SMPL
L1060-18 - 1	TP-3	144	-0.0266	-	7470A	PPB	10 Jan 2020 18:20:04	U	SMPL
PB126015BL - 1	PBW	220	-0.0110	-	7470A	PPB	10 Jan 2020 18:22:10	U	SMPL
PB126015BS - 1	LCSW	18197	3.6837	-	7470A	PPB	10 Jan 2020 18:24:16	U	SMPL
K6484-02 - 1	40121	232	-0.0085	-	7470A	PPB	10 Jan 2020 18:26:22	U	SMPL
K6484-04 - 1	30627	147	-0.0260	-	7470A	PPB	10 Jan 2020 18:28:28	U	SMPL
K6484-06 - 1	30628	420	0.0301	-	7470A	PPB	10 Jan 2020 18:30:35	U	SMPL
L1047-02 - 1	FMI109	97	-0.0362	-	7470A	PPB	10 Jan 2020 18:32:42	U	SMPL
L1047-02DUP - 1	FMI109DUP	146	-0.0262	-	7470A	PPB	10 Jan 2020 18:34:49	U	SMPL
L1047-02MS - 1	FMI109MS	15445	3.1181	-	7470A	PPB	10 Jan 2020 18:36:57	U	SMPL
L1047-02MSD - 1	FMI109MSD	15988	3.2297	-	7470A	PPB	10 Jan 2020 18:39:03	U	SMPL
CCV42 - 1	CCV42	23287	4.7298	-	7470A	PPB	10 Jan 2020 18:41:09	U	SMPL
CCB42 - 1	CCB42	251	-0.0046	-	7470A	PPB	10 Jan 2020 18:43:15	U	SMPL
PB126023BL - 1	PB126023BL	211	-0.0128	-	7470A	PPB	10 Jan 2020 18:45:22	U	SMPL
PB126023BS - 1	PB126023BS	16356	3.3053	-	7470A	PPB	10 Jan 2020 18:47:28	U	SMPL
L1064-03 - 1	SOIL-A	117	-0.0321	-	7470A	PPB	10 Jan 2020 18:49:34	U	SMPL
L1064-07 - 1	SOIL-B	235	-0.0079	-	7470A	PPB	10 Jan 2020 18:51:40	U	SMPL
L1064-10 - 1	OD	297	0.0049	-	7470A	PPB	10 Jan 2020 18:53:47	U	SMPL
L1065-01 - 1	6-HER-002-003-01	287	0.0028	-	7470A	PPB	10 Jan 2020 18:55:54	U	SMPL
L1070-03 - 1	RBR-200030	276	0.0005	-	7470A	PPB	10 Jan 2020 18:58:01	U	SMPL
L1070-06 - 1	RBR-200059	248	-0.0052	-	7470A	PPB	10 Jan 2020 19:00:08	U	SMPL
L1084-04 - 1	AF-1	348	0.0153	-	7470A	PPB	10 Jan 2020 19:02:16	U	SMPL
L1084-05 - 1	AF-2	323	0.0102	-	7470A	PPB	10 Jan 2020 19:04:22	U	SMPL
CCV43 - 1	CCV43	23388	4.7505	-	7470A	PPB	10 Jan 2020 19:06:28	U	SMPL
CCB43 - 1	CCB43	218	-0.0114	-	7470A	PPB	10 Jan 2020 19:08:34	U	SMPL
L1084-06 - 1	AF-3	295	0.0044	-	7470A	PPB	10 Jan 2020 19:10:40	U	SMPL
L1084-06DUP - 1	AF-3DUP	289	0.0032	-	7470A	PPB	10 Jan 2020 19:12:46	U	SMPL
L1084-06MS - 1	AF-3MS	16731	3.3824	-	7470A	PPB	10 Jan 2020 19:14:52	U	SMPL
L1084-06MSD - 1	AF-3MSD	16450	3.3246	-	7470A	PPB	10 Jan 2020 19:16:58	U	SMPL
PB126014TB - 1	PB125962TB	192	-0.0167	-	7470A	PPB	10 Jan 2020 19:19:05	U	SMPL
PB126015TB - 1	PB125965TB	226	-0.0097	-	7470A	PPB	10 Jan 2020 19:21:12	U	SMPL
PB126023TB - 1	PB125985TB	226	-0.0097	-	7470A	PPB	10 Jan 2020 19:23:18	U	SMPL
L1060-05LX5 - 1	L1060-05LX5	41	-0.0478	-	7470A	PPB	10 Jan 2020 19:25:26	U	SMPL
L1047-02LX5 - 1	L1047-02LX5	34	-0.0492	-	7470A	PPB	10 Jan 2020 19:27:34	U	SMPL
L1084-06LX5 - 1	L1084-06LX5	65	-0.0428	-	7470A	PPB	10 Jan 2020 19:29:41	U	SMPL
CCV44 - 1	CCV44	23250	4.7222	-	7470A	PPB	10 Jan 2020 19:31:47	U	SMPL
CCB44 - 1	CCB44	249	-0.0050	-	7470A	PPB	10 Jan 2020 19:33:53	U	SMPL

01/10/20