



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB107132

Analyst By : JIGNESH
Supervisor Review By : apatel
Slope : 98.5
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/09/2020	15:30
2	CAL2	1	Water	NA	NA	20.1	7.01	01/09/2020	15:31
3	CAL3	1	Water	NA	NA	20.1	10.03	01/09/2020	15:32
4	ICV	1	Water	NA	NA	20.1	7.00	01/09/2020	15:34
5	CCV1	1	Water	NA	NA	20.1	2.01	01/09/2020	15:35
6	L1065-02	1	Water	NA	NA	18.1	7.17	01/09/2020	15:36
7	L1074-01	1	Water	NA	NA	12.1	6.31	01/09/2020	15:37
8	L1074-02	1	Water	NA	NA	13.4	6.44	01/09/2020	15:38
9	L1074-02DUP	1	Water	NA	NA	13.5	6.45	01/09/2020	15:39
10	CCV2	1	Water	NA	NA	20.2	12.02	01/09/2020	15:40

MS 107132

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1065 W WorkList ID : 134909 Department : Wet-Chemistry Date : 01-09-2020 15:12:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/08/2020	Water	L1065-02	Corrosivity	Cool 4 deg C	OBR103	I13	6-HER-002-003-02	01/08/2020	9040C
01/08/2020	Water	L1074-01	Corrosivity	Cool 4 deg C	PARS15	I21	CIY1-GW-02-13.5	01/08/2020	9040C
01/08/2020	Water	L1074-02	Corrosivity	Cool 4 deg C	PARS15	I21	CIY1-GW-102-13.5	01/08/2020	9040C

Date/Time 01-09-2020 15:20
 Received by: JG
 Relinquished by: DK

Date/Time 01-09-2020 16:20
 Received by: DK
 Relinquished by: JG