



PERCENT SOLID

Supervisor: apatel
Analyst: Daniel
Date: 1/14/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 01/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:57
Out Date: 01/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107137

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1004-01	BU-RO-258	14	1.12	9.71	8.99	91.6	
L1004-02	BU-RO-258	15	1.14	9.54	8.61	88.9	
L1004-04	BU-RO-234	16	1.16	9.97	9.53	95.0	
L1004-05	BU-RO-234	17	1.15	9.78	8.73	87.8	
L1004-07	BU-RO-229	18	1.13	9.71	8.58	86.8	
L1004-08	BU-RO-229	19	1.17	9.72	8.64	87.4	
L1082-05	WASTE	20	1.18	9.93	8.77	86.7	
L1103-01	SB-MW-14-(2-2.5)	1	1.14	9.95	7.06	67.2	
L1103-02	SB-15-(2-3.5)	2	1.13	9.8	9.03	91.1	
L1103-03	SB-MW-16-(2-3.5)	3	1.16	9.92	8.96	89.0	
L1103-04	SB-17-(2-4)	4	1.19	9.86	8.78	87.5	
L1103-05	DUP-01-01092020	5	1.15	9.71	8.73	88.6	
L1103-06	SB-18-(3-5)	6	1.13	9.54	8.98	93.3	
L1103-07	SB-MW-19-(3-5)	7	1.16	9.63	9.11	93.9	
L1103-08	L1103-07MS	8	1.16	9.63	9.11	93.9	
L1103-09	L1103-07MSD	9	1.16	9.63	9.11	93.9	
L1103-10	SB-1-(4-6)	10	1.18	9.76	9.32	94.9	
L1103-11	SB-20-(3-5)	11	1.11	9.94	9.09	90.4	
L1107-01	BW-2-3.0	12	1.18	9.64	8.45	85.9	
L1107-02	BW-209-3.0	13	1.18	9.79	8.3	82.7	
L1108-01	INWOOD-BH-01-A	24	1.15	9.93	8.61	85.0	
L1108-02	INWOOD-BH-01-B	25	1.18	9.71	8.46	85.3	
L1108-03	INWOOD-BH-02-A	26	1.17	9.87	8.48	84.0	
L1108-04	INWOOD-BH-02-B	27	1.18	9.6	9.05	93.5	
L1108-05	INWOOD-BH-03-A	28	1.13	9.54	8.2	84.1	
L1108-06	INWOOD-BH-03-B	29	1.1	9.85	8.8	88.0	
L1111-01	OUTSIDE-BRICK	21	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS
L1111-04	OUTSIDE-FOUNDATIONS	22	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS



PERCENT SOLID

Supervisor: apatel
Analyst: Daniel
Date: 1/14/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 01/13/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:57
Out Date: 01/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107137

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1111-07	CONCRETE-FLOOR	23	1.00	2.00	2.00	100.0	STONE SAMPLE 100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$