


LB 107149

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	202	0.000	202				
0.2	0.200	0.248	0.048	1601	0.0 %	1601				
2.5	2.500	2.531	0.031	13317	0.0 %	13317				
5.0	5.000	4.940	-0.060	25677	0.0 %	25677				
7.5	7.500	7.442	-0.058	38517	0.0 %	38517				
10.0	10.000	10.063	0.063	51967	0.0 %	51967				

LB107149 INSTRUMENT ID :CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jan 2020 16:09:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD-0	202	-	0.0000	7470A	PPB	13 Jan 2020 16:11:25	S	Std
0.2 - 1	SD-2	1601	-	0.2000	7470A	PPB	13 Jan 2020 16:13:32	S	Std
2.5 - 1	SD-5	13317	-	2.5000	7470A	PPB	13 Jan 2020 16:15:40	S	Std
5.0 - 1	SD-10	25677	-	5.0000	7470A	PPB	13 Jan 2020 16:17:48	S	Std
7.5 - 1	SD-15	38517	-	7.5000	7470A	PPB	13 Jan 2020 16:19:55	S	Std
10.0 - 1	SD-20	51967	-	10.0000	7470A	PPB	13 Jan 2020 16:22:01	S	Std
ICV67 - 1	ICV67	19745	3.7839	-	7470A	PPB	13 Jan 2020 16:25:55	U	SMPL
ICB67 - 1	ICB67	1153	0.1608	-	7470A	PPB	13 Jan 2020 16:28:00	U	SMPL
CCV45 - 1	CCV45	27278	5.2519	-	7470A	PPB	13 Jan 2020 16:30:05	U	SMPL
CCB45 - 1	CCB45	824	0.0967	-	7470A	PPB	13 Jan 2020 16:32:11	U	SMPL
CRA - 1	CRA	1441	0.2169	-	7470A	PPB	13 Jan 2020 16:37:33	U	SMPL
HighStd - 1	HighStd	51715	10.0141	-	7470A	PPB	13 Jan 2020 16:39:38	U	SMPL
ChkStd - 1	ChkStd	36181	6.9869	-	7470A	PPB	13 Jan 2020 16:41:44	U	SMPL
PB126053BL - 1	PBW	1164	0.1629	-	7470A	PPB	13 Jan 2020 16:43:50	U	SMPL
PB126053BS - 1	LCSW	17111	3.2706	-	7470A	PPB	13 Jan 2020 16:45:57	U	SMPL
L1047-01 - 1	FMI109	315	-0.0025	-	7470A	PPB	13 Jan 2020 16:48:03	U	SMPL
L1065-02 - 1	6-HER-002-003-02	624	0.0577	-	7470A	PPB	13 Jan 2020 16:50:10	U	SMPL
L1065-02DUP - 1	6-HER-002-003-02DUP	655	0.0637	-	7470A	PPB	13 Jan 2020 16:52:18	U	SMPL
L1065-02MS - 1	6-HER-002-003-02MS	17275	3.3026	-	7470A	PPB	13 Jan 2020 16:54:26	U	SMPL
L1065-02MSD - 1	6-HER-002-003-02MSD	17141	3.2765	-	7470A	PPB	13 Jan 2020 16:56:31	U	SMPL
CCV46 - 1	CCV46	25924	4.9880	-	7470A	PPB	13 Jan 2020 16:58:37	U	SMPL
CCB46 - 1	CCB46	-8	-0.0655	-	7470A	PPB	13 Jan 2020 17:00:43	U	SMPL
L1074-01 - 1	CIY1-GW-02-13.5	214	-0.0222	-	7470A	PPB	13 Jan 2020 17:02:49	U	SMPL
L1074-02 - 1	CIY1-GW-102-13.5	135	-0.0376	-	7470A	PPB	13 Jan 2020 17:04:55	U	SMPL
PB126068BL - 1	PBW	34	-0.0573	-	7470A	PPB	13 Jan 2020 17:07:00	U	SMPL
PB126068BS - 1	LCSW	17079	3.2644	-	7470A	PPB	13 Jan 2020 17:09:07	U	SMPL
L1014-05 - 1	OR-03-011020	278	-0.0097	-	7470A	PPB	13 Jan 2020 17:11:13	U	SMPL
L1014-05DUP - 1	OR-03-011020DUP	301	-0.0053	-	7470A	PPB	13 Jan 2020 17:13:20	U	SMPL
L1014-05MS - 1	OR-03-011020MS	20804	3.9903	-	7470A	PPB	13 Jan 2020 17:20:13	U	SMPL
L1014-05MSD - 1	OR-03-011020MSD	20739	3.9776	-	7470A	PPB	13 Jan 2020 17:22:20	U	SMPL
L1091-02 - 1	GRAVEL-SOIL	287	-0.0080	-	7470A	PPB	13 Jan 2020 17:24:27	U	SMPL
L1106-03 - 1	CIY1-WC-SB-BC-04	376	0.0094	-	7470A	PPB	13 Jan 2020 17:26:34	U	SMPL
CCV47 - 1	CCV47	26644	5.1284	-	7470A	PPB	13 Jan 2020 17:28:40	U	SMPL
CCB47 - 1	CCB47	200	-0.0249	-	7470A	PPB	13 Jan 2020 17:30:46	U	SMPL
L1106-05 - 1	CIY1-WC-SB-02-03	305	-0.0045	-	7470A	PPB	13 Jan 2020 17:32:52	U	SMPL
L1106-07 - 1	CIY1-WC-SB-01	295	-0.0064	-	7470A	PPB	13 Jan 2020 17:34:59	U	SMPL
L1065-02LX5 - 1		141	-0.0364	-	7470A	PPB	13 Jan 2020 17:37:04	U	SMPL
L1014-05LX5 - 1		65	-0.0513	-	7470A	PPB	13 Jan 2020 17:39:11	U	SMPL
PB126053TB - 1	PB126025TB	135	-0.0376	-	7470A	PPB	13 Jan 2020 17:41:17	U	SMPL
CCV48 - 1	CCV48	26126	5.0274	-	7470A	PPB	13 Jan 2020 17:43:24	U	SMPL
CCB48 - 1	CCB48	8	-0.0624	-	7470A	PPB	13 Jan 2020 17:45:31	U	SMPL

BP

01/13/20