

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB107160

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP81171
calibration std. phosphate 0.5 ppm	WP81170
calibration std. phosphate 0.3 ppm	WP81169
calibration std. phosphate 0.1 ppm	WP81168
calibration std. phosphate 0.05 ppm	WP81167
calibration std. 0 ppm	WP81166
phosphate ICV std.	WP81173
phosphate CCV std.	WP81172
5N sulfuric acid	WP79628
Combined reagent	WP81175
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314

Intercept: -0.0003Slope: 0.7075Regression: 0.999938

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	01/14/2020	13:17
2	CAL2	0.05	1	50	50	0.037	01/14/2020	13:17
3	CAL3	0.10	1	50	50	0.072	01/14/2020	13:18
4	CAL4	0.30	1	50	50	0.211	01/14/2020	13:18
5	CAL5	0.50	1	50	50	0.348	01/14/2020	13:19
6	CAL6	1.00	1	50	50	0.710	01/14/2020	13:19



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB107160

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.330	0.467	01/14/2020	13:20
2	ICB		1	50	50	0.000	0.000	01/14/2020	13:20
3	CCV1	0.50	1	50	50	0.336	0.475	01/14/2020	13:21
4	CCB1		1	50	50	0.000	0.000	01/14/2020	13:21
5	PB126083BL		1	50	50	0.000	0.000	01/14/2020	13:22
6	PB126083BS	0.50	1	50	50	0.329	0.465	01/14/2020	13:22
7	L1112-01		1	50	50	0.013	0.019	01/14/2020	13:23
8	L1112-02		1	50	50	0.016	0.023	01/14/2020	13:23
9	L1112-02DUP		1	50	50	0.017	0.024	01/14/2020	13:24
10	L1112-02MS	0.50	1	50	50	0.331	0.468	01/14/2020	13:24
11	L1112-02MSD	0.50	1	50	50	0.331	0.468	01/14/2020	13:25
12	CCV2	0.50	1	50	50	0.336	0.475	01/14/2020	13:26
13	CCB2		1	50	50	0.000	0.000	01/14/2020	13:26