



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 1/15/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:33
In Date: 01/14/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/15/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB107165

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L1005-01	HR-05-011420	1	1.16	9.69	8.84	90.0	
L1005-02	HR-05-011420	2	1.16	9.69	9.00	91.9	
L1005-03	HR-05-011420-1	3	1.15	9.95	9.23	91.8	
L1005-04	HR-05-011420-2	4	1.16	9.84	8.71	87.0	
L1114-01	RBR-200057-27	5	1.14	9.83	8.79	88.0	
L1115-01	SIDE-WALK-COMP	6	1.00	2.00	2.00	100.0	Concrete Sample
L1123-01	SB-01-5.5-6.0	7	1.15	9.5	7.5	76.0	
L1123-02	SB-01-COMP	8	1.13	9.81	8.14	80.8	
L1123-03	SB-02-2.5-3.0	9	1.14	9.58	8.31	85.0	
L1123-04	SB-02-COMP	10	1.12	9.74	7.9	78.7	
L1123-05	SB-03-5.5-6.0	11	1.12	9.98	8.34	81.5	
L1123-06	SB-03-COMP	12	1.12	9.77	7.79	77.1	
L1123-07	SB-04-2.5-3.0	13	1.14	9.58	8.26	84.4	
L1123-08	SB-04-COMP	14	1.13	9.98	8.77	86.3	
L1125-01	FESB-28- (1.5-2)	15	1.12	9.96	8.19	80.0	
L1125-02	FESB-25A- (6.5-7)	16	1.12	9.62	8.09	82.0	
L1125-03	FESB-34- (4.5-5)	17	1.14	9.56	8.09	82.5	
L1125-04	FESB-21A- (2.5-3)	18	1.15	9.96	8.15	79.5	
L1125-05	FESB-33- (3.5-4)	19	1.13	9.74	7.94	79.1	
L1127-01	QVD-SB-08-1.5-2	20	1.13	9.85	9.00	90.3	
L1127-02	QVD-SB-07-1.5-2	21	1.14	9.65	8.91	91.3	
L1127-03	QVD-SB-02-2-2.5	22	1.14	9.7	7.7	76.6	
L1127-04	QVD-SB-102-2-2.5	23	1.15	9.65	7.73	77.4	
L1127-05	QVD-SB-03-1.5-2	24	1.15	9.78	8.8	88.6	
L1127-06	QVD-SB-01-1.5-2	25	1.15	9.57	8.43	86.5	
L1128-01	391231708.0	26	1.15	9.98	2.84	19.1	SLUDGE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$