



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 1/15/2020

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:33  
In Date: 01/14/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN #1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:00  
Out Date: 01/15/2020  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB107165

| Lab ID   | Client SampleID   | Dish # | Dish Wt (g) (A) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments        |
|----------|-------------------|--------|-----------------|--------------------------|----------------------------|---------|-----------------|
| L1005-01 | HR-05-011420      | 1      | 1.16            | 9.69                     | 8.84                       | 90.0    |                 |
| L1005-02 | HR-05-011420      | 2      | 1.16            | 9.69                     | 9.00                       | 91.9    |                 |
| L1005-03 | HR-05-011420-1    | 3      | 1.15            | 9.95                     | 9.23                       | 91.8    |                 |
| L1005-04 | HR-05-011420-2    | 4      | 1.16            | 9.84                     | 8.71                       | 87.0    |                 |
| L1114-01 | RBR-200057-27     | 5      | 1.14            | 9.83                     | 8.79                       | 88.0    |                 |
| L1115-01 | SIDE-WALK-COMP    | 6      | 1.00            | 2.00                     | 2.00                       | 100.0   | Concrete Sample |
| L1123-01 | SB-01-5.5-6.0     | 7      | 1.15            | 9.5                      | 7.5                        | 76.0    |                 |
| L1123-02 | SB-01-COMP        | 8      | 1.13            | 9.81                     | 8.14                       | 80.8    |                 |
| L1123-03 | SB-02-2.5-3.0     | 9      | 1.14            | 9.58                     | 8.31                       | 85.0    |                 |
| L1123-04 | SB-02-COMP        | 10     | 1.12            | 9.74                     | 7.9                        | 78.7    |                 |
| L1123-05 | SB-03-5.5-6.0     | 11     | 1.12            | 9.98                     | 8.34                       | 81.5    |                 |
| L1123-06 | SB-03-COMP        | 12     | 1.12            | 9.77                     | 7.79                       | 77.1    |                 |
| L1123-07 | SB-04-2.5-3.0     | 13     | 1.14            | 9.58                     | 8.26                       | 84.4    |                 |
| L1123-08 | SB-04-COMP        | 14     | 1.13            | 9.98                     | 8.77                       | 86.3    |                 |
| L1125-01 | FESB-28- (1.5-2)  | 15     | 1.12            | 9.96                     | 8.19                       | 80.0    |                 |
| L1125-02 | FESB-25A- (6.5-7) | 16     | 1.12            | 9.62                     | 8.09                       | 82.0    |                 |
| L1125-03 | FESB-34- (4.5-5)  | 17     | 1.14            | 9.56                     | 8.09                       | 82.5    |                 |
| L1125-04 | FESB-21A- (2.5-3) | 18     | 1.15            | 9.96                     | 8.15                       | 79.5    |                 |
| L1125-05 | FESB-33- (3.5-4)  | 19     | 1.13            | 9.74                     | 7.94                       | 79.1    |                 |
| L1127-01 | QVD-SB-08-1.5-2   | 20     | 1.13            | 9.85                     | 9.00                       | 90.3    |                 |
| L1127-02 | QVD-SB-07-1.5-2   | 21     | 1.14            | 9.65                     | 8.91                       | 91.3    |                 |
| L1127-03 | QVD-SB-02-2-2.5   | 22     | 1.14            | 9.7                      | 7.7                        | 76.6    |                 |
| L1127-04 | QVD-SB-102-2-2.5  | 23     | 1.15            | 9.65                     | 7.73                       | 77.4    |                 |
| L1127-05 | QVD-SB-03-1.5-2   | 24     | 1.15            | 9.78                     | 8.8                        | 88.6    |                 |
| L1127-06 | QVD-SB-01-1.5-2   | 25     | 1.15            | 9.57                     | 8.43                       | 86.5    |                 |
| L1128-01 | 391231708.0       | 26     | 1.15            | 9.98                     | 2.84                       | 19.1    | SLUDGE SAMPLE   |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

LB 10-2-165

WorkList Name : %1-011420      WorkList ID : 135009      Department : Wet-Chemistry      Date : 01-14-2020 09:16:24

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample  | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|------------------|--------------|--------------|
| 01/21/2020 | Solid  | L1005-01 | Percent Solids | Cool 4 deg C | PSEG05   | I22              | HR-05-011420     | 01/14/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1005-02 | Percent Solids | Cool 4 deg C | PSEG05   | I22              | HR-05-011420     | 01/14/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1005-03 | Percent Solids | Cool 4 deg C | PSEG05   | I22              | HR-05-011420-1   | 01/14/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1005-04 | Percent Solids | Cool 4 deg C | PSEG05   | I22              | HR-05-011420-2   | 01/14/2020   | Chemtech -SO |
| 01/20/2020 | Solid  | L1114-01 | Percent Solids | Cool 4 deg C | PSEG01   | I23              | RBR-200057-27    | 01/14/2020   | Chemtech -SO |
| 01/17/2020 | Solid  | L1115-01 | Percent Solids | Cool 4 deg C | PSEG01   | I23              | SIDE-WALK-COMP   | 01/14/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-01 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-01-5.5-6.0    | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-02 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-01-COMP       | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-03 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-02-2.5-3.0    | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-04 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-02-COMP       | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-05 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-03-5.5-6.0    | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-06 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-03-COMP       | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-07 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-04-2.5-3.0    | 01/13/2020   | Chemtech -SO |
| 01/21/2020 | Solid  | L1123-08 | Percent Solids | Cool 4 deg C | LIRO01   | I13              | SB-04-COMP       | 01/13/2020   | Chemtech -SO |
| 01/23/2020 | Solid  | L1125-01 | Percent Solids | Cool 4 deg C | FIRS02   | I22              | FESB-28-(1.5-2)  | 01/13/2020   | Chemtech -SO |
| 01/24/2020 | Solid  | L1125-02 | Percent Solids | Cool 4 deg C | FIRS02   | I22              | FESB-25A-(6.5-7) | 01/14/2020   | Chemtech -SO |
| 01/24/2020 | Solid  | L1125-03 | Percent Solids | Cool 4 deg C | FIRS02   | I22              | FESB-34-(4.5-5)  | 01/14/2020   | Chemtech -SO |
| 01/24/2020 | Solid  | L1125-04 | Percent Solids | Cool 4 deg C | FIRS02   | I22              | FESB-21A-(2.5-3) | 01/14/2020   | Chemtech -SO |
| 01/24/2020 | Solid  | L1125-05 | Percent Solids | Cool 4 deg C | FIRS02   | I22              | FESB-33-(3.5-4)  | 01/14/2020   | Chemtech -SO |
| 01/23/2020 | Solid  | L1127-01 | Percent Solids | Cool 4 deg C | PARS15   | I22              | QVD-SB-08-1.5-2  | 01/13/2020   | Chemtech -SO |
| 01/23/2020 | Solid  | L1127-02 | Percent Solids | Cool 4 deg C | PARS15   | I22              | QVD-SB-07-1.5-2  | 01/13/2020   | Chemtech -SO |
| 01/23/2020 | Solid  | L1127-03 | Percent Solids | Cool 4 deg C | PARS15   | I22              | QVD-SB-02-2.2-5  | 01/13/2020   | Chemtech -SO |

Date/Time 01/14/2020 17:00  
 Received by: DK  
 Relinquished by: CB

Date/Time 01/14/2020  
 Received by: CB  
 Relinquished by: DK

CBS107165

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-011420      WorkList ID : 135009      Department : Wet-Chemistry      Date : 01-14-2020 09:16:24

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Storage Location | Customer Sample  | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|------------------|------------------|--------------|--------------|
| 01/23/2020 | Solid  | L1127-04 | Percent Solids | Cool 4 deg C | PARS15   | I22              | QVD-SB-102-2-2.5 | 01/13/2020   | Chemtech -SO |
| 01/24/2020 | Solid  | L1127-05 | Percent Solids | Cool 4 deg C | PARS15   | I22              | QVD-SB-03-1.5-2  | 01/14/2020   | Chemtech -SO |
| 01/24/2020 | Solid  | L1127-06 | Percent Solids | Cool 4 deg C | PARS15   | I22              | QVD-SB-01-1.5-2  | 01/14/2020   | Chemtech -SO |
| 01/05/2020 | Solid  | L1128-01 | Percent Solids | Cool 4 deg C | GARD04   | I22              | 391231708.01     | 12/26/2019   | Chemtech -SO |

Date/Time 01/14/2020  
Received by: PK  
Relinquished by: [Signature]

Date/Time 01/14/2020  
Received by: [Signature]  
Relinquished by: PK