



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107177

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/15/2020	10:20
2	L1123-13		1	8	100.00	>212.0	>212.0	01/15/2020	10:45
3	L1123-13DUP		1	8	100.00	>212.0	>212.0	01/15/2020	11:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB107177

WorkList Name : fp11520 WorkList ID : 135041 Department : Wet-Chemistry Date : 01-15-2020 09:19:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/21/2020	Water	L1123-13	Flash Point	Cool 4 deg C	LIRO01	I13	TWP-01	01/14/2020	1010A

Date/Time 01/15/20 09:25
 Received by: KS
 Relinquished by: DK

Date/Time 01/15/20 11:30
 Received by: DK
 Relinquished by: KS