

LB107178


7471B

INSTRUMENT ID: CVI

Linear

 μ Abs.:

4714E



A= 0.0000e+000

B= 2.1158e-004

C= -9.0100e-002

Rho= 0.9997745

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.072	-0.072	87	0.000	87				
0.2	0.200	0.234	0.034	1531	0.0 %	1531				
2.5	2.500	2.474	-0.026	12119	0.0 %	12119				
5.0	5.000	5.068	0.068	24381	0.0 %	24381				
7.5	7.500	7.610	0.110	36393	0.0 %	36393				
10.0	10.000	9.885	-0.115	47148	0.0 %	47148				

LB107178 INSTRUMENT ID:CV1

Method: 7471B		Operator: Admin		Date of Analysis: 15 Jan 2020 14:32:23					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	87	-	0.0000	7471B	PPB	15 Jan 2020 14:58:50	S	Std
0.2 - 1	50.2	1531	-	0.2000	7471B	PPB	15 Jan 2020 15:01:07	S	Std
2.5 - 1	52.5	12119	-	2.5000	7471B	PPB	15 Jan 2020 15:03:24	S	Std
5.0 - 1	55.0	24381	-	5.0000	7471B	PPB	15 Jan 2020 15:05:40	S	Std
7.5 - 1	57.5	36393	-	7.5000	7471B	PPB	15 Jan 2020 15:07:57	S	Std
10.0 - 1	50.0	47148	-	10.0000	7471B	PPB	15 Jan 2020 15:34:30	S	Std
ICV69 - 1	ICV69	21130	4.3806	-	7471B	PPB	15 Jan 2020 15:37:31	U	SMPL
ICB69 - 1	ICB69	183	-0.0514	-	7471B	PPB	15 Jan 2020 15:39:45	U	SMPL
CCV54 - 1	CCV54	25313	5.2656	-	7471B	PPB	15 Jan 2020 15:46:22	U	SMPL
CCB54 - 1	CCB54	184	-0.0512	-	7471B	PPB	15 Jan 2020 15:48:36	U	SMPL
CRA - 1	CRA	1586	0.2455	-	7471B	PPB	15 Jan 2020 15:50:51	U	SMPL
HighStd - 1	HighStd	50921	10.6838	-	7471B	PPB	15 Jan 2020 15:57:05	U	SMPL
ChkStd - 1	ChkStd	32959	6.8834	-	7471B	PPB	15 Jan 2020 16:03:41	U	SMPL
PB126092BL - 1	PBS	153	-0.0577	-	7471B	PPB	15 Jan 2020 16:05:56	U	SMPL
PB126092BS - 1	LCSS	21772	4.5164	-	7471B	PPB	15 Jan 2020 16:08:12	U	SMPL
L1004-01 - 1	BU-RO-258	2744	0.4905	-	7471B	PPB	15 Jan 2020 16:10:28	U	SMPL
L1004-01DUP - 1	BU-RO-258DUP	2897	0.5228	-	7471B	PPB	15 Jan 2020 16:12:46	U	SMPL
L1004-01MS - 1	BU-RO-258MS	24154	5.0204	-	7471B	PPB	15 Jan 2020 16:20:21	U	SMPL
L1004-01MSD - 1	BU-RO-258MSD	23829	4.9516	-	7471B	PPB	15 Jan 2020 16:24:52	U	SMPL
L1004-04 - 1	BU-RO-234	2682	0.4774	-	7471B	PPB	15 Jan 2020 16:27:07	U	SMPL
CCV55 - 1	CCV55	25461	5.2969	-	7471B	PPB	15 Jan 2020 16:29:22	U	SMPL
CCB55 - 1	CCB55	123	-0.0641	-	7471B	PPB	15 Jan 2020 16:31:37	U	SMPL
L1004-07 - 1	BU-RO-229	4040	0.7647	-	7471B	PPB	15 Jan 2020 16:33:54	U	SMPL
L1108-01 - 1	INWOOD-BH-01-A	13653	2.7986	-	7471B	PPB	15 Jan 2020 16:36:09	U	SMPL
L1108-02 - 1	INWOOD-BH-01-B	12867	2.6323	-	7471B	PPB	15 Jan 2020 16:38:25	U	SMPL
L1108-03 - 1	INWOOD-BH-02-A	18697	3.8658	-	7471B	PPB	15 Jan 2020 16:40:41	U	SMPL
L1108-04 - 1	INWOOD-BH-02-B	6638	1.3144	-	7471B	PPB	15 Jan 2020 16:42:56	U	SMPL
L1108-05 - 1	INWOOD-BH-03-A	15775	3.2476	-	7471B	PPB	15 Jan 2020 16:45:12	U	SMPL
L1108-06 - 1	INWOOD-BH-03-B	11643	2.3733	-	7471B	PPB	15 Jan 2020 16:47:28	U	SMPL
L1114-01 - 1	RBR-200057-27	1635	0.2558	-	7471B	PPB	15 Jan 2020 16:49:45	U	SMPL
L1005-01 - 1	HR-05-011420	4777	0.9206	-	7471B	PPB	15 Jan 2020 16:52:01	U	SMPL
L1125-01 - 1	FESB-28-(1.5-2)	2295	0.3955	-	7471B	PPB	15 Jan 2020 16:54:17	U	SMPL
CCV56 - 1	CCV56	25226	5.2472	-	7471B	PPB	15 Jan 2020 16:56:32	U	SMPL
CCB56 - 1	CCB56	154	-0.0575	-	7471B	PPB	15 Jan 2020 16:58:47	U	SMPL
L1125-02 - 1	FESB-25A-(6.5-7)	1804	0.2916	-	7471B	PPB	15 Jan 2020 17:01:03	U	SMPL
L1125-03 - 1	FESB-34-(4.5-5)	4473	0.8563	-	7471B	PPB	15 Jan 2020 17:03:19	U	SMPL
L1125-04 - 1	FESB-21A-(2.5-3)	1359	0.1974	-	7471B	PPB	15 Jan 2020 17:05:34	U	SMPL
L1125-05 - 1	FESB-33-(3.5-4)	958	0.1126	-	7471B	PPB	15 Jan 2020 17:07:50	U	SMPL
L1126-01 - 1	BW-2-20.0	419	-0.0014	-	7471B	PPB	15 Jan 2020 17:10:06	U	SMPL
L1126-02 - 1	BW-2-45.0	453	0.0057	-	7471B	PPB	15 Jan 2020 17:12:22	U	SMPL
L1126-03 - 1	BW-1-3.0	2745	0.4907	-	7471B	PPB	15 Jan 2020 17:14:38	U	SMPL
L1126-04 - 1	L1126-03MS	24189	5.0278	-	7471B	PPB	15 Jan 2020 17:16:55	U	SMPL
L1126-05 - 1	L1126-03MSD	24040	4.9963	-	7471B	PPB	15 Jan 2020 17:19:12	U	SMPL
CCV57 - 1	CCV57	22884	4.7517	-	7471B	PPB	15 Jan 2020 17:21:27	U	SMPL
CCB57 - 1	CCB57	158	-0.0567	-	7471B	PPB	15 Jan 2020 17:23:43	U	SMPL
PB126114BL - 1	PBS	317	-0.0230	-	7471B	PPB	15 Jan 2020 17:25:59	U	SMPL
PB126114BS - 1	LCSS	21172	4.3895	-	7471B	PPB	15 Jan 2020 17:35:14	U	SMPL
L1126-03DUP - 1		2704	0.4820	-	7471B	PPB	15 Jan 2020 17:37:29	U	SMPL
L1126-03LX5 - 1		492	0.0140	-	7471B	PPB	15 Jan 2020 17:39:45	U	SMPL
L1004-01LX5 - 1		518	0.0195	-	7471B	PPB	15 Jan 2020 17:42:00	U	SMPL
CCV58 - 1	CCV58	23335	4.8471	-	7471B	PPB	15 Jan 2020 17:44:15	U	SMPL
CCB58 - 1	CCB58	284	-0.0300	-	7471B	PPB	15 Jan 2020 17:46:32	U	SMPL