



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107185

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/15/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/15/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/15/2020	16:02
4	ICV	1	Water	NA	NA	20.1	7.00	01/15/2020	16:05
5	CCV1	1	Water	NA	NA	20.1	2.01	01/15/2020	16:06
6	L1145-01	1	Water	NA	NA	14.4	7.18	01/15/2020	16:10
7	L1145-01DUP	1	Water	NA	NA	14.5	7.19	01/15/2020	16:11
8	L1145-02	1	Water	NA	NA	20.6	6.01	01/15/2020	16:12
9	CCV2	1	Water	NA	NA	20.1	12.02	01/15/2020	16:15

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1145 WorkList ID : 135060 Department : Wet-Chemistry Date : 01-15-2020 15:36:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2020	Water	L1145-01 -D	pH	Cool 4 deg C	TULL02	I31	MH-5	01/15/2020	9040C
01/15/2020	Water	L1145-02 -D	pH	Cool 4 deg C	TULL02	I31	MH-109	01/15/2020	9040C

Date/Time 01-15-2020 1540
 Received by: Jq
 Relinquished by:

Date/Time 01-15-2020 1645
 Received by: Dk
 Relinquished by: Jq