



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107186

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	01/15/2020	16:00
2	CAL2	1	Water	NA	NA	20.1	7.01	01/15/2020	16:01
3	CAL3	1	Water	NA	NA	20.1	10.03	01/15/2020	16:02
4	ICV	1	Water	NA	NA	20.1	7.00	01/15/2020	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/15/2020	17:25
6	L1006-04	1	Solid	20.03	20	21.5	8.46	01/15/2020	17:30
7	L1006-04DUP	1	Solid	20.04	20	21.6	8.47	01/15/2020	17:32
8	L1016-09	1	Solid	20.02	20	21.2	8.62	01/15/2020	17:35
9	L1017-06	1	Solid	20.03	20	21.2	8.78	01/15/2020	17:40
10	L1126-06	1	Solid	20.03	20	22.1	9.13	01/15/2020	17:42
11	CCV2	1	Water	NA	NA	20.1	12.01	01/15/2020	17:45

YB-107186

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1016			WorkList ID : 135057		Department : Wet-Chemistry		Date : 01-15-2020 14:17:34		
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2020	Solid	L1006-04	Corrosivity	Cool 4 deg C	PSEG05	I23	HD-01-011520	01/15/2020	9045D
01/15/2020	Solid	L1016-09	Corrosivity	Cool 4 deg C	PSEG05		SU-03-011520	01/15/2020	9045D
01/15/2020	Solid	L1017-06	Corrosivity	Cool 4 deg C	PSEG05		TR-05-011520	01/15/2020	9045D
01/14/2020	Solid	L1126-06	Corrosivity	Cool 4 deg C	CDMS01	I41	WASTE-CHARACTERIZATION	01/14/2020	9045D

Date/Time 01-15-2020 16:15
Received by: J.
Relinquished by: DL

Date/Time 01-15-2020 18:10
Received by: DL
Relinquished by: J.