



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107191

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/16/2020	09:35
2	L1142-02		1	8	100.00	>212.0	>212.0	01/16/2020	09:55
3	L1142-02DUP		1	8	100.00	>212.0	>212.0	01/16/2020	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 107191

WorkList Name : fp11620

WorkList ID : 135087

Department : Wet-Chemistry

Date : 01-16-2020 08:45:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/28/2020	Water	L1142-02	Flash Point	Cool 4 deg C	ATCE02	I31	WS-1	01/15/2020	1010A

Date/Time 01/16/20 09:00
 Received by: KS
 Relinquished by: DK

Date/Time 01/16/20 10:25
 Received by: DK
 Relinquished by: KS