



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107200

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/16/2020	11:00
2	CAL2	1	Water	NA	NA	20.2	7.00	01/16/2020	11:01
3	CAL3	1	Water	NA	NA	20.2	10.03	01/16/2020	11:02
4	ICV	1	Water	NA	NA	20.1	7.01	01/16/2020	11:05
5	CCV1	1	Water	NA	NA	20.2	2.01	01/16/2020	11:06
6	L1151-01	1	Water	NA	NA	17.7	5.21	01/16/2020	11:08
7	L1151-01DUP	1	Water	NA	NA	17.8	5.22	01/16/2020	11:09
8	CCV2	1	Water	NA	NA	20.1	12.02	01/16/2020	11:10

WORKLIST(Hardcopy Internal Chain)

VB 107200

WorkList Name : PH L1151

WorkList ID : 135095

Department : Wet-Chemistry

Date : 01-16-2020 10:44:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/10/2020	Water	L1151-01	pH	Cool 4 deg C	STER10	I23	CELL-A -POOL-WATER(DE-IC)	01/10/2020	SM 4500-H B

Date/Time 01-16-2020 10:45
 Received by: SO
 Relinquished by: SO

Date/Time 01-16-2020 12:00
 Received by: SO
 Relinquished by: SO