



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107202

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP81256
Turbidity Calibration std, 0NTU	WP81255
Turbidity Calibration std, 20NTU	WP81247
Turbidity Calibration std, 40NTU	WP81246
Turbidity Calibration std, 1NTU	WP81250
Turbidity Calibration std, 10NTU	WP81248
Turbidity Calibration std, 5NTU	WP81249
10 NTU Standard 500 ml	W2489

Intercept: -0.1609

Slope: 0.9175

Regression: 0.9994

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.136	01/16/2020	11:35
2	CAL2	1	1	0.991	01/16/2020	11:37
3	CAL3	5	1	4.390	01/16/2020	11:40
4	CAL4	10	1	8.950	01/16/2020	11:42
5	CAL5	20	1	17.300	01/16/2020	11:45
6	CAL6	40	1	37.000	01/16/2020	11:49



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB107202

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.260	10.268	01/16/2020	11:50
2	ICB		1	0.130	0.317	01/16/2020	11:53
3	CCV1	10	1	8.990	9.974	01/16/2020	11:55
4	CCB1		1	0.135	0.323	01/16/2020	11:58
5	LB107202BL		1	0.130	0.317	01/16/2020	12:01
6	L1134-01		1	3.430	3.914	01/16/2020	12:07
7	CCV2	10	1	8.960	9.941	01/16/2020	12:10
8	CCB2		1	0.132	0.319	01/16/2020	12:13

LB107202 WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity01162020 WorkList ID : 135090 Department : Wet-Chemistry Date : 01-16-2020 09:13:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/15/2020	Water	L1134-01	Turbidity	Cool 4 deg C	CHEM02	QA Office	PT-TURB-WS	01/13/2020	180.1

Date/Time 01-16-2020
 Received by: De
 Relinquished by: Ad

Date/Time 01-16-2020
 Received by: NA
 Relinquished by: NA