



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB107211

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/16/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	01/16/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	01/16/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.00	01/16/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	01/16/2020	17:16
6	L1008-14	1	Solid	20.03	20	20.6	8.75	01/16/2020	17:19
7	L1008-14DUP	1	Solid	20.04	20	20.7	8.76	01/16/2020	17:20
8	L1010-09	1	Solid	20.02	20	20.9	7.27	01/16/2020	17:22
9	L1012-06	1	Solid	20.03	20	20.6	8.12	01/16/2020	17:25
10	L1154-02	1	Solid	20.04	20	20.7	8.51	01/16/2020	17:30
11	L1154-04	1	Solid	20.02	20	20.8	8.66	01/16/2020	17:32
12	L1154-06	1	Solid	20.03	20	21.1	8.83	01/16/2020	17:35
13	L1154-08	1	Solid	20.04	20	20.8	8.48	01/16/2020	17:36
14	L1155-03	1	Solid	20.02	20	21.2	9.44	01/16/2020	17:39
15	L1155-06	1	Solid	20.01	20	21.2	6.97	01/16/2020	17:40
16	CCV2	1	Water	NA	NA	20.1	12.02	01/16/2020	17:42

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L1155

WorkList ID : 135111

Department : Wet-Chemistry

Date : 01-16-2020 14:44:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2020	Solid	L1008-14	Corrosivity	Cool 4 deg C	PSEG05	I31	CL-01-011620	01/16/2020	9045D
01/16/2020	Solid	L1010-09	Corrosivity	Cool 4 deg C	PSEG05	I31	OK-02-011620	01/16/2020	9045D
01/16/2020	Solid	L1012-06	Corrosivity	Cool 4 deg C	PSEG05	I31	EO-02-011620	01/16/2020	9045D
01/16/2020	Solid	L1154-02	Corrosivity	Cool 4 deg C	PSEG01	I31	B4-0-15	01/16/2020	9045D
01/16/2020	Solid	L1154-04	Corrosivity	Cool 4 deg C	PSEG01	I31	B4-15-20	01/16/2020	9045D
01/16/2020	Solid	L1154-06	Corrosivity	Cool 4 deg C	PSEG01	I31	B3-0-15	01/16/2020	9045D
01/16/2020	Solid	L1154-08	Corrosivity	Cool 4 deg C	PSEG01	I31	B3-15-20	01/16/2020	9045D
01/16/2020	Solid	L1155-03	Corrosivity	Cool 4 deg C	PSEG01	I31	RBR-200028	01/16/2020	9045D
01/16/2020	Solid	L1155-06	Corrosivity	Cool 4 deg C	PSEG01	I31	RBR-200052	01/16/2020	9045D

Date/Time

Received by:

Relinquished by:

01-16-2020 16:00

Jo

Jo

Date/Time

Received by:

Relinquished by:

01-16-2020 17:30

Jo

Jo