



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107212

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/16/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.00	01/16/2020	17:11
3	CAL3	1	Water	NA	NA	20.1	10.03	01/16/2020	17:12
4	ICV	1	Water	NA	NA	20.1	7.00	01/16/2020	17:15
5	CCV1	1	Water	NA	NA	20.1	2.01	01/16/2020	17:16
6	L1162-01	1	Water	NA	NA	14.1	7.36	01/16/2020	17:45
7	L1162-02	1	Water	NA	NA	14.2	7.38	01/16/2020	17:46
8	L1162-03	1	Water	NA	NA	14.0	7.41	01/16/2020	17:48
9	L1162-04	1	Water	NA	NA	14.8	6.76	01/16/2020	17:50
10	L1162-05	1	Water	NA	NA	14.1	7.65	01/16/2020	17:52
11	L1162-08	1	Water	NA	NA	14.9	7.33	01/16/2020	18:00
12	L1162-09	1	Water	NA	NA	15.3	7.02	01/16/2020	18:02
13	L1162-10	1	Water	NA	NA	16.1	6.85	01/16/2020	18:05
14	L1162-10DUP	1	Water	NA	NA	16.2	6.86	01/16/2020	18:06
15	CCV2	1	Water	NA	NA	20.1	12.01	01/16/2020	18:08

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L1162 WorkList ID : 135112 Department : Wet-Chemistry Date : 01-16-2020 16:55:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/16/2020	Water	L1162-01 -C	pH	Cool 4 deg C	TULL02	I31	AOC-9-NORTH	01/16/2020	9040C
01/16/2020	Water	L1162-02 -C	pH	Cool 4 deg C	TULL02	I31	CBH-591	01/16/2020	9040C
01/16/2020	Water	L1162-03 -C	pH	Cool 4 deg C	TULL02	I31	MH-5-(SETTLED)	01/16/2020	9040C
01/16/2020	Water	L1162-04 -C	pH	Cool 4 deg C	TULL02	I31	CBH-594	01/16/2020	9040C
01/16/2020	Water	L1162-05 -C	pH	Cool 4 deg C	TULL02	I31	MH-109-(SETTLED)	01/16/2020	9040C
01/16/2020	Water	L1162-08 -C	pH	Cool 4 deg C	TULL02	I31	CBH-140	01/16/2020	9040C
01/16/2020	Water	L1162-09 -C	pH	Cool 4 deg C	TULL02	I31	MH-719	01/16/2020	9040C
01/16/2020	Water	L1162-10 -C	pH	Cool 4 deg C	TULL02	I31	MH-720	01/16/2020	9040C

Date/Time 01/16/2020 17:00
 Received by: SA
 Relinquished by: SA

Date/Time 01/16/2020 18:15
 Received by: AK
 Relinquished by: SA