



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107223

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER, COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/17/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.00	01/17/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	01/17/2020	11:19
4	ICV	1	Water	NA	NA	20.1	7.00	01/17/2020	11:20
5	CCV1	1	Water	NA	NA	20.1	2.01	01/17/2020	11:22
6	L1174-01	1	Water	NA	NA	16.5	4.67	01/17/2020	11:24
7	L1174-01DUP	1	Water	NA	NA	16.6	4.68	01/17/2020	11:25
8	CCV2	1	Water	NA	NA	20.1	12.01	01/17/2020	11:26