



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB107233

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	01/17/2020	15:10
2	CAL2	1	Water	NA	NA	20.1	7.01	01/17/2020	15:11
3	CAL3	1	Water	NA	NA	20.1	10.03	01/17/2020	15:12
4	ICV	1	Water	NA	NA	20.2	7.00	01/17/2020	15:15
5	CCV1	1	Water	NA	NA	20.1	2.01	01/17/2020	15:16
6	L1166-01	1	Water	NA	NA	16.8	6.60	01/17/2020	15:19
7	L1166-01DUP	1	Water	NA	NA	16.9	6.61	01/17/2020	15:20
8	L1184-01	1	Water	NA	NA	16.5	7.22	01/17/2020	15:22
9	L1184-02	1	Water	NA	NA	16.6	6.98	01/17/2020	15:24
10	L1184-03	1	Water	NA	NA	17.3	7.51	01/17/2020	15:25
11	L1184-04	1	Water	NA	NA	17.7	6.92	01/17/2020	15:26
12	L1184-05	1	Water	NA	NA	17.9	6.95	01/17/2020	15:30
13	L1184-06	1	Water	NA	NA	17.2	7.61	01/17/2020	15:32
14	L1184-07	1	Water	NA	NA	12.3	7.41	01/17/2020	15:35
15	L1184-08	1	Water	NA	NA	14.0	7.58	01/17/2020	15:36
16	CCV2	1	Water	NA	NA	20.2	12.01	01/17/2020	15:40