



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB107251

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	01/20/2020	09:40
2	L1166-01		1	8	100.00	>212.0	>212.0	01/20/2020	10:10
3	L1166-01DUP		1	8	100.00	>212.0	>212.0	01/20/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 107251

WorkList Name : fp12020 WorkList ID : 135172 Department : Wet-Chemistry Date : 01-20-2020 09:00:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
01/29/2020	Water	L1166-01	Flash Point	Cool 4 deg C	PARS15	I33	QVD-WC-L-01	01/15/2020	1010A

Date/Time 01/20/20 09:05
 Received by: KS
 Relinquished by: DK

Date/Time 01/20/20 10:50
 Received by: DK
 Relinquished by: KS